# Point Implant

Good implants. Be happy world!

Ver. 15.1



www.pointimplant.com

# Point Implant

## Good implants. Be happy world!

## By World 1st UV Implant Technology!



SM I
Fixture
**Tissue Level**

SM I
Wide neck Fixture
**Tissue Level**

SM II
Short Fixture
**Bone Level**


SM II
Fixture
**Bone Level**

TN
Fixture
**Bone Level**

# Point Implant
# Innovative R&D

## We Lead the Implant Technology!

Lead the dentistry by developing innovative products to become the pioneer in dental field with the state of the art UV technology.




## Point Implant Slogan
## Good implants. Be happy world!

Good implants are essential for the success of clinics and the satisfaction of patients.
Our mission is to provide the world's best quality.

Point Implant
Good implants. Be happy world!

## Color System
## Royal Blue


Point Implant
Good implants. Be happy world!

PANTONE 286 C

Point Implant
Good implants. Be happy world!

**CMYK** C 100 M 6 Y 0 K 2
**RGB** R 0 G 56 B 168
**HEX** #0038A8

Implants will be firmly rooted
thanks to **our innovative UV technology.**



UV activation without breaking
the sterilized condition of the fixtures.

# History for Last 32 Years

| | | |
|---|---|---|
| 2026 | | MDR CE To be done |
| 2025 | | Fixture 400k shipped to China, 1st World symposium in Bangkok, Launched Quantum CBCT |
| 2024 | • SEP | Obtained implant certificate from China and India |
| 2023 | • APR | Obtained USA FDA 510k of UV Implant, Company name changed from 'PointNix' to 'Point Implant' |
| 2022 | • JUN | Expansion and relocation of HQ office |
| | • JAN | Obtained CE(MDD) certificate of UV Implant |
| 2020 | • DEC | Obtained FDA 510k for Point 500/500C |
| 2019 | • DEC | Acquired BNB Systems for laser |
| 2017 | • JAN | Acquired Hanyang for unit chair |
| 2016 | • JAN | Launched MyCeph S/W & UV Activator |
| 2015 | • NOV | Launched UV Active Implant |
| | • FEB | Selected as 'Global Small Giant Company' |
| 2014 | • DEC | Established subsidiary in China |
| | • APR | Development of UV Implant started |
| 2012 | • DEC | Acquired Hallim Dentech Co., Ltd |
| 2011 | • DEC | Released CDX-View mobile version |
| | • FEB | Obtained CE certificate for Point 500C HD / 500 HD |
| 2008 | • OCT | 1st launching of implant system |
| 2007 | • NOV | Obtained CE certificate for Dental CT / Panorama |
| 2006 | • SEP | Launched 1st Dental CT / Panorama model |
| 2001 | • AUG | Established own R&D center |
| 1998 | • AUG | Company name changed to PointNix Co. after merging Nix Software Co. for ENT application |
| 1993 | • JAN | Established Point System Co. producing intra-oral camera |

# Contents


## 01 Implant INTRO

SM I 07
SM II 08
TN 09
UV Activator 11
Initial stability of UV Implant 13
ISQ 15
Process Smart Factory 17

## 02 SM II

**SM II UV Fixture (No-Mount)** 22
**Short UV Fixture** 23
**SM II SLA Fixture (No-Mount)** 25
Cover Screw 26
Healing Abutment
Bite Impression Coping Scanbody 27
Bite Impression Coping
SM II Scanbody 28
Fixture Pick-up Impression Coping
Fixture Transfer Impression Coping 29
Fixture Lab Analog
Ball Lab Analog 30
Cemented Abutment
Angled Abutment 31
Milling Abutment 32
Temporary Abutment
CCM Casting Abutment 33
Solid Abutment 34
Solid Outer Driver 35
Loc Abutment
Ball Abutment 36
Ball Abutment Driver
Solid Protect Cap

## 03 TN

**TN Deep Thread Blade Type Fixture** 37

## 04 SM I

**SM I UV Fixture (No-Mount)** 40
**SM I SLA Fixture (No-Mount)** 41
Cover Screw
Healing Abutment 40
SM I Scanbody
Fixture Pick-up Impression Coping
Fixture Transfer Impression Coping
Solid Lab Analog 43
Excellent Solid Lab Analog
Cemented Abutment
Excellent Cemented Abutment 44
Angled Abutment
Milling Abutment 45
Excellent Temporary Abutment
CCM Casting Abutment
Solid Abutment 46
Excellent Solid Abutment
Fixture Pin
Excellent Solid Abutment

## 05 Digital Prosthetics

Bite Impression Coping Scanbody 49
SM II Scanbody
Customized Abutment
Ti-Base 50
Digital 3D Analog
Cemented Abutment 51



## 06 MUA


Multi Straight Abutment 54
Multi Angled Abutment
Multi Abutment Driver 55
Multi Healing Cap
Multi Scan body
Multi Tranfer Impression Coping
Multi Lab Analog
Multi Impression Coping 56
Ti-Base MUA
Multi Combination Cylinder
Multi Temporary Cylinder
Multi Casting Cylinder 57
Multi Plastic Cylinder
Ti-Base Screw
Multi Cylinder Screw


## 07 KIT


Point Surgical Full Kit (OTF-KIT) 59
Point Surgical Standard Kit (OTA-KIT) 60
Point Surgical Compact Kit (OTS-KIT) 61
Point NAVI Kit 67
Point Narrow NAVI Kit 71
Drilling protocol 75
Point Prosthetic Kit 77
Bite Impression Kit 79
Point Full S.S.T KIT 80
Point Simple S.S.T KIT 81
Point S.O.S KIT 82
Bone Expander KIT 83
Sinus Crestal Approach KIT
Sinus Lateral Approach KIT 84
Implant & Screw Remover KIT(s)
Simple Screw Remover KIT


## 08 Clical Case Report 85

## 09 Other Products


Point NAVI 97
Dental CBCT 101
Dental Unit Chair 103
Dental Laser 105

# SM I

## Implant Design

Tissue Level



Regular ( G / H : 1.8 ) Regular ( G / H : 2.8 ) Wide ( G / H : 2.0 ) Wide ( G / H : 3.0 )

### Crestal bone protection

Interface is positioned above bone to reduce crestal remodeling.

### Smooth Collar (Tissue-Level)

Interface is positioned above bone to reduce crestal remodeling.

### S.L.A Surface

S.L.A surface with optimal roughness. Enlarged bone contact area for good osseointegration.



### Stable soft tissue

Tissue-level neck for consistent soft-tissue thickness.

### 8° Octa Connection

Tapered octa design for strong, stable, microgap-safe connection.

### Progressive Thread

Enhanced bone contact and excellent primary stability.

# SM II

## Implant Design

Bone Level



Mini | Regular | Ultra-Wide | Short

### Prevent crestal bone loss

Designed to redude stress around fixture and prevent crestal bone resorption.

### S.L.A Surface

S.L.A surface with optimal roughness. Enlarged bone contact area for good osseointegration.

### Wide Cutting Edge

Minimize pressure on surrounding bone.



### Reduce the risk of fracture

Increased thickness to strengthen connection area and prevent fixture fracture.

### Optimal thread pitch Wing thread


Increased bone contact area.

### Root form design

Fast implant osseointegration and excellent initial stability Flat ended design to minimize nerve damage and sinus perforation.

# TN

Beyond Limits.
Redefining Excellent Implant.
Deep Thread Blade Type.



Point Implant system

# TN

## Implant Design

Bone Level / Deep Thread Blade Type



Mini Regular Ultra-Wide

### Superior performance in soft bone(D3, D4)

Deep, aggressive threads secure strong anchorage in weak bone, maximizing fixation where it is typically difficult.

### S.L.A Surface

S.L.A surface with optimal roughness. Enlarged bone contact area for good osseointegration.

### Progressive core for controlled compression

Progressive core improves bone engagement and controlled compression.



### Preserves cortical bone

Minimizes removal of the cortical plate, supporting crestal bone preservation.

### Optimized for immediate loading


Reliable stability for immediate placement and loading.

### High-performance cutting performance cutting installation

Aggressive threads with cutting flutes allow precise control and smooth placement.

# UV Activator

UV irradiation equipment for fixtures.



# UV Activator

## Implant UV

Irradiation device

UV Activator irradiates implant surface with ultravio et light to improve adhesion of organic matter. It converts surface charge, and activates super-hydrophilic surface.

**12 Hours**
Guarantee Hydrocarbon removal and Rapid. Osseointegration within 12 hours of activation.

**Up to 36 Fixtures**
Capability of 36 implants Simultaneous activation.

**1 Year Warranty**
High Standard Quality control ensure Products to be free from defect in material and workmanship.

**UV-C**
Light spectrum with specific wavelength inactivate more than 99.9% of viruses and bacteria.

### Photocatalytic **UV Implant Effect**



**Super Hydrophilic**
Excellent bioactivity



**Anti-aging effect**
Surface aging prevention (Hydrocarbon removal).



**Maximize cell activity**
Surface charge recovery.





**Bacterial adhesion inhibitory effect**
Surface antibacterial properties maintenance.



**Induction of rapid bone formation**
Contact Osteogenesis.


### Product **Characteristics**





| | |
|---|---|
| **Elctrical rating** | 220V, 60Hz, 90W |
| **Dimension** | h : 399mm w : 400mm d : 493mm |
| **Grade of insulation** | Class I |
| **Operatimg Priciple** | This device irradiates UV light through a UV lamp and removes a hydrocarbon film on an aging implant fixture, there by increasing hydrophilicity, reducing the time to restore and heal during an implant surgery. |

# Initial stability of UV Implant




1. The Superior bioactivity of UV Implant cramatically increases secondary stability and removes the stability dip.
2. The UV Implant at 2weeks after placement shows the excellent osseointegration corresponding to general implant in weeks 8.
3. The ISQ value of the UV Implant next generation of innovative is increased from the first week, and osseointegraion is completed in 3weeks. The initial implant failure rate is close to zero.
4. UV Implant has an excellent osseointegration rate of 98%, which is excellent for long-term safety compared to general implants with 50% to 60% osseointegration.

## Comparison of **Implant Blood Activity**



Point UV Implant | A company | B company

Titanium surfaces that have been in constant time after implant production (eg, more than one month of surface treatment) are hydrophobic.
That is, the water contact angle is higher than 90 degrees. Which is close to 90 degrees in most surface types.
Hydrophobicity is common in the surface morphology of all titanium and appears in the form of hemisphere that is not absorbed by the surface of the fixture.
Very interestingly, after UV treatment these surfaces are surprisingly well wetted, resulting in a zero contact angle, which is called superhydrophilic.
The superhydrophilic surface has a positive effect on the BIC of the implants, the rate of osseointegration, and various blood factors involved in bone formation.

## | UV Implant Optimized for Early Loading

### Start Prosthetic Process in just 2 weeks!

Osstell equipment is used to evaluate implant stability and the degree of osseointegration. The measured value, ranging from 1 to 100, is referred to as the Implant Stability Quotient (ISQ).
Generally, at the time of implant placement, ISQ values range from 55 to 75 in the maxilla and from 65 to 85 in the mandible.
The mechanical primary stability formed at the initial stage of implant placement gradually decreases, while secondary stability resulting from bone regeneration and remodeling gradually increases. During this transition period, a phenomenon known as the stability dip occurs.
As demonstrated in numerous studies, ISQ values typically decrease until approximately three weeks after implant placement, fluctuate between six and eight weeks, and then gradually increase thereafter. Consequently, it has been reported that prosthetic loading should be avoided around three weeks after implant placement.
While many implant manufacturers have focused on improving initial stability through thread design, Point Implant has developed a breakthrough next-generation surface treatment technology through continuous research, enabling rapid osseointegration and significantly increasing the bone adhesion rate.
As a result, prosthetic loading can be performed as early as two weeks after implant placement.
Furthermore, the osseointegration rate of UV-treated implants at two weeks was equivalent to that of conventional SLA implants at eight weeks, and the bone adhesion rate of UV implants reached 98% at four weeks, compared to 53% for conventional implants.

## | Change in ISQ value after a week of placement (UV Implant)

Case 1




On the day of placement.




In a week after placement.

Case 2




On the day of placement.




In a week after placement.

Case 3




On the day of placement.




In a week after placement.

Case 4




On the day of placement.




In a week after placement.

Case 5




On the day of placement.



In a week after placement.

Case 6



On the day of placement.



In a week after placement.


# ISQ

## Change in ISQ value of UV Implant

Point Implant R&D Center has been researching and developing UV implants since 2010. Through extensive research and clinical trials conducted over the past seven years, Point Implant developed the world's first capsule-type UV implant, demonstrating osteointegration equivalent to that of a conventional SLA implant at eight weeks.
Objective assessments of implant stability and osseointegration were performed using Osstell equipment. As a result, ISQ values showed a significant increase from the first week after placement.
Due to the early increase in ISQ values, the elimination of the stability dip, and the shortened osseointegration period, the UV Implant enables stable immediate placement and early loading.

## Significant increases of ISQ value during 4 weeks

(Total Case 1241 : Max Imp 582, Mand Imp 659)




Over the last 10 years, Point R&D Center with Dr. Kyou-Hwa Park from Seoul national University measured ISQ value from the day of surgery to 4th week after placement of UV Implant in various clinical case.
ISQ value after surgery was increased in almost every clinical cases.



Fig. A : Contact Osteogenesis



Fig. B : Distance Osteogenesis

We Lead the Implant Technology!

# Point UV Implant

Lead the dentistry by developing innovative products
to become the pioneer in dental field with the state of the art technology.



# Process
# Smart Factory

Quality control target in the Implant factory: below 100PPM defects.
(Less than 1 out of 10,000 pieces)

## CNC Production Process.




## High Pressure Steam Automatic Cleaning Equipment.




## Fixture Inspection Automation Equipment.




## CS HA Automatic Inspection Machine.




## Abutment Screw Inspection Automation Equipment.




## Fixture (UV) Packaging Automation Line.




## Blister & Box Packaging Automation Line.




For Optimal Product Selection Across All Clinical Cases.

# Point Implant Line-up

## SM I




## SM I
Wide neck




## SM II




## SM II
Short




Connection: 1–2 mm selectable.

## TN




# SM II



**SM II UV Fixture (No-Mount)**
**Short UV Fixture**
**SM II SLA Fixture (No-Mount)**
Cover Screw
Healing Abutment
Bite Impression Coping Scanbody
Bite Impression Coping
SM II Scanbody
Fixture Pick-up Impression Coping
Fixture Transfer Impression Coping
Fixture Lab Analog
Ball Lab Analog
Cemented Abutment
Angled Abutment
Milling Abutment
Temporary Abutment
CCM Casting Abutment
Solid Abutment
Solid Outer Driver
Loc Abutment
Ball Abutment
Ball Abutment Driver
Solid Protect Cap

# SM II Flowchart



# SM II UV Fixture (No-Mount)



- Submerged Type implant with Internal Hex "I" Taper Connection Structure.
- Optimal thread design.
- Taper body design with excellent initial stability.
- Strong Self Threading Effect with Corkscrew design.
- Possible to acquire the initial stability for immediate loading even on soft bones.
- Surface : SLA Surface + UV.
  (Effect of maximizing hydrophilicity by expanding the UV irradiation absorption area)
- Recommended mounting torque: 40Ncm or less.
- It is recommended to use a fixture with a diameter of 4.5mm or more for the posterior single case.

## UV Fixture







- **Mini**

| D \ L | 8.5 | 10 | 11.5 | 13 |
|---|---|---|---|---|
| Ø 3.0 | POF3008QNP | POF3010QNP | POF3011QNP | POF3013QNP |
| Ø 3.5 | POF3508QNP | POF3510QNP | POF3511QNP | POF3513QNP |
| Ø 4.0 | POF4008MQNP | POF4010MQNP | POF4011MQNP | POF4013MQNP |

- **Regular**

| D \ L | 7 | 8.5 | 10 | 11.5 | 13 |
|---|---|---|---|---|---|
| Ø 4.0 | POF4007QNP | POF4008QNP | POF4010QNP | POF4011QNP | POF4013QNP |
| Ø 4.5 | POF4507QNP | POF4508QNP | POF4510QNP | POF4511QNP | POF4513QNP |
| Ø 5.0 | POF5007QNP | POF5008QNP | POF5010QNP | POF5011QNP | POF5013QNP |
| Ø 5.5 | POF5507QNP | POF5508QNP | POF5510QNP | POF5511QNP | POF5513QNP |

- **Ultra-Wide**

| D \ L | 7 | 8.5 | 10 | 11.5 | 13 |
|---|---|---|---|---|---|
| Ø 6.0 | POF6007QNP | POF6008QNP | POF6010QNP | POF6011QNP | POF6013QNP |
| Ø 6.5 | POF6507QNP | POF6508QNP | POF6510QNP | POF6511QNP | POF6513QNP |
| Ø 7.0 | POF7007QNP | POF7008QNP | POF7010QNP | POF7011QNP | POF7013QNP |

## SM II UV Fixture
(Pre-Mount)



- Submerged-type implant with internal hex and "I" taper connection
- Optimized taper body and corkscrew thread for high primary stability
- Stable immediate loading, even in soft bone
- SLA + UV surface for enhanced osseointegration
- Recommended insertion torque: ≤ 40 Ncm
- Ø 4.5 mm or larger fixtures for single molar cases

- **Mini**

| D \ L | 8.5 | 10 | 11.5 | 13 |
|---|---|---|---|---|
| Ø 3.0 | POF3008Q | POF3010Q | POF3011Q | POF3013Q |
| Ø 3.5 | POF3508Q | POF3510Q | POF3511Q | POF3513Q |

- **Regular**

| D \ L | 7 | 8.5 | 10 | 11.5 | 13 |
|---|---|---|---|---|---|
| Ø 4.0 | POF4007Q | POF4008Q | POF4010Q | POF4011Q | POF4013Q |
| Ø 4.5 | POF4507Q | POF4508Q | POF4510Q | POF4511Q | POF4513Q |
| Ø 5.0 | POF5007Q | POF5008Q | POF5010Q | POF5011Q | POF5013Q |
| Ø 5.5 | POF5507Q | POF5508Q | POF5510Q | POF5511Q | POF5513Q |

- **Ultra-Wide**

| D \ L | 7 | 8.5 | 10 | 11.5 | 13 |
|---|---|---|---|---|---|
| Ø 6.0 | POF6007Q | POF6008Q | POF6010Q | POF6011Q | POF6013Q |
| Ø 6.5 | POF6507Q | POF6508Q | POF6510Q | POF6511Q | POF6513Q |
| Ø 7.0 | POF7007Q | POF7008Q | POF7010Q | POF7011Q | POF7013Q |

# Short

## | Short UV Fixture




### Short Key Advantages

1. **Implantable even in insufficient bone**
   Can be implanted directly in the maxilla or mandible without additional bone grafting even when bone height is inadequate. → Particularly beneficial for elderly patients or cases where sinus lift is difficult or undesired.

2. **Reduced surgical burden**
   Allows omission of major procedures such as bone grafting and sinus lift, leading to shorter surgery time, reduced bleeding, and faster recovery.


3. **Enhanced patient safety and convenience**
   Reduced bone grafting lowers the risk of complications such as infection, swelling, and pain. It is a practical option for elderly patients or those with systemic diseases.




* The Bevel has a Clean Surface.

• **Regular**

| D \ B | 1.0 | 2.0 |
|---|---|---|
| Ø 4.0 | POF400710NP | POF400720NP |
| Ø 4.5 | POF450710NP | POF450720NP |
| Ø 5.0 | POF500710NP | POF500720NP |
| Ø 5.5 | POF550710NP | POF550720NP |
| Ø 6.0 | POF600710NP | POF600720NP |
| Ø 6.5 | POF650710NP | POF650720NP |
| Ø 7.0 | POF700710NP | POF700720NP |




1. L1 (Connection Height): 1-2mm → Soft tissue. Enhances prosthetic stability.
2. L2 (Thread Length): 5-6mm → Immediately applicable even in thin bone.
3. F4.0-F7.0 Optimal Design → Improved load distribution, increased initial stability, and ease of procedure.
4. Overall Length: 7mm structure→ Excellent accessibility.
5. Regular Platform Only.

Point Implant
Good implants. Be happy world!


## Dual platform (Regular, Mini)

# For Ø4.0 fixtures

to provide optimal solutions for various cases.



# SM II SLA Fixture (No-Mount)




- Submerged Type Implant with Internal Hex 11° Taper Connection Structure.
- Optimal thread design.
- Taper body design with excellent initial stability.
- Strong Self Threading Effect with Corkscrew design.
- Possible to acquire the initial stability for immediate loading even on soft bones.
- Surface : SLA.
- Recommended mounting torque: 40Ncm or less
- It is recommended to use a fixture with a diameter of 4.5mm or more for the posterior single case.

## SLA Fixture



- **Mini**

| D \ L | 8.5 | 10 | 11.5 | 13 |
|---|---|---|---|---|
| Ø 3.0 | POF3008 | POF3010 | POF3011 | POF3013 |
| Ø 3.5 | POF3508 | POF3510 | POF3511 | POF3513 |
| Ø 4.0 | POF4008M | POF4010M | POF4011M | POF4013M |



- **Regular**

| D \ L | 7 | 8.5 | 10 | 11.5 | 13 |
|---|---|---|---|---|---|
| Ø 4.0 | POF4007 | POF4008 | POF4010 | POF4011 | POF4013 |
| Ø 4.5 | POF4507 | POF4508 | POF4510 | POF4511 | POF4513 |
| Ø 5.0 | POF5007 | POF5008 | POF5010 | POF5011 | POF5013 |
| Ø 5.5 | POF5507 | POF5508 | POF5510 | POF5511 | POF5513 |



- **Ultra-Wide**

| D \ L | 7 | 8.5 | 10 | 11.5 | 13 |
|---|---|---|---|---|---|
| Ø 6.0 | POF6007 | POF6008 | POF6010 | POF6011 | POF6013 |
| Ø 6.5 | POF6507 | POF6508 | POF6510 | POF6511 | POF6513 |
| Ø 7.0 | POF7007 | POF7008 | POF7010 | POF7011 | POF7013 |

## Cover Screw



OCS4000G OCS4001GA OCS4002GA

- Packing Unit : Cover Screw.
- 1.2 Hex Driver / with 10 N of torque.

- **Mini**

| D \ G/H | - | 1.0 | 2.0 |
|---|---|---|---|
| Ø 3.0 | OCS3000GA | OCS3001GA | OCS3002GA |
| Ø 3.5 | OCS3500GA | OCS3501GA | OCS3502GA |

- **Regular**

| D \ G/H | - | 1.0 | 2.0 |
|---|---|---|---|
| Ø 4.0 | OCS4000C | OCS4001GA | OCS4002GA |

## Healing Abutment



G / H

- Packing Unit : Healing Abutment.
- 1.2 Hex Driver / with 10 N of torque.
- Re-tighten with a torque of 30Ncm after 10 minutes of initial placement.

- **Mini**

| D \ G/H | 2.0 | 3.0 | 4.0 | 5.0 | 6.0 | 7.0 | 8.0 | 9.0 |
|---|---|---|---|---|---|---|---|---|
| Ø 3.5 | OSHAM3502SG | OSHAM3503SG | OSHAM3504SG | OSHAM3505SG | OSHAM3506SG | OSHAM3507SG | OSHAM3508SG | OSHAM3509SG |
| Ø 4.0 | - | OSHAM4003SG | OSHAM4004SG | OSHAM4005SG | OSHAM4006SG | OSHAM4007SG | OSHAM4008SG | OSHAM4009SG |
| Ø 4.5 | - | OSHAM4503SG | OSHAM4504SG | OSHAM4505SG | OSHAM4506SG | OSHAM4507SG | OSHAM4508SG | OSHAM4509SG |
| Ø 5.0 | - | OSHAM5003SG | OSHAM5004SG | OSHAM5005SG | OSHAM5006SG | OSHAM5007SG | OSHAM5008SG | OSHAM5009SG |

- **Regular**

| D \ G/H | 2.0 | 3.0 | 4.0 | 5.0 | 6.0 | 7.0 | 8.0 | 9.0 |
|---|---|---|---|---|---|---|---|---|
| Ø 4.0 | OSHA4002SG | OSHA4003SG | OSHA4004SG | OSHA4005SG | OSHA4006SG | OSHA4007SG | OSHA4008SG | OSHA4009SG |
| Ø 4.5 | OSHA4502SG | OSHA4503SG | OSHA4504SG | OSHA4505SG | OSHA4506SG | OSHA4507SG | OSHA4508SG | OSHA4509SG |
| Ø 5.0 | OSHA5002SG | OSHA5003SG | OSHA5004SG | OSHA5005SG | OSHA5006SG | OSHA5007SG | OSHA5008SG | OSHA5009SG |
| Ø 5.5 | OSHA5502SG | OSHA5503SG | OSHA5504SG | OSHA5505SG | OSHA5506SG | OSHA5507SG | OSHA5508SG | OSHA5509SG |
| Ø 6.0 | OSHA6002SG | OSHA6003SG | OSHA6004SG | OSHA6005SG | OSHA6006SG | OSHA6007SG | OSHA6008SG | OSHA6009SG |
| Ø 6.5 | OSHA6502SG | OSHA6503SG | OSHA6504SG | OSHA6505SG | OSHA6506SG | OSHA6507SG | OSHA6508SG | OSHA6509SG |
| Ø 7.0 | OSHA7002SG | OSHA7003SG | OSHA7004SG | OSHA7005SG | OSHA7006SG | OSHA7007SG | OSHA7008SG | OSHA7009SG |

## Bite Impression Coping Scanbody




- Packing Unit : Scan body + Guide Pin.
- Scan Body for Titanium Custom Abutment Fabrication.
- For Model Scan : Long (15mm).
- For Intra-Oral Scan : Short (10mm).
- Fastening by hand using 1.2 Hex Driver.

**Size Classification**

- Guide Pin : Regular(Non), Mini(Brown).
- Body : G/H 2mm(Non), 3mm(Yellow), 4mm(Purple), 6mm(Blue), 8mm(Non) Coloring.

**• Mini**

| D TYPE G/H | 2.0 | 3.0 | 4.0 | 6.0 | 8.0 |
|---|---|---|---|---|---|
| Ø 4.0 Scan body | OBICM4002SBP | OBICM4003SBP | OBICM4004SBP | OBICM4006SBP | OBICM4008SBP |

**• Regular**

| D TYPE G/H | 2.0 | 3.0 | 4.0 | 6.0 | 8.0 |
|---|---|---|---|---|---|
| Ø 4.0 Scan body | OBICR4002SBP | OBICR4003SBP | OBICR4004SBP | OBICR4006SBP | OBICR4008SBP |
| Ø 4.5 Scan body | OBICR4502SBP | OBICR4503SBP | OBICR4504SBP | OBICR4506SBP | OBICR4508SBP |
| Ø 5.0 Scan body | OBICR5002SBP | OBICR5003SBP | OBICR5004SBP | OBICR5006SBP | OBICR5008SBP |

## Bite Impression Coping




- Packing Unit : Bite Impression Coping Body + Guide Pin.
- Components for taking fixture level impressions.
- Bite impression can be taken at the same time as impression taking.
- Confirmation of implant connection by checking screw and post height.
- A library that can also be used as a scan body.
- Cuter Driver : Short(OBICDSP), Long(OBICDLP).

**Size Classification**

- Guide Pin : Regular(Non), Mini(Brown).
- Body : G/H 2mm(Non), 3mm(Yellow), 4mm(Purple), 6mm(Blue), 8mm(Non) Coloring.

**• Mini**

| D TYPE G/H | | 2.0 | 3.0 | 4.0 | 6.0 | 8.0 |
|---|---|---|---|---|---|---|
| Ø 4.0 | Hex | OBICM4002P | OBICM4003P | OBICM4004P | OBICM4006P | OBICM4008P |
| Ø 4.5 | Hex | OBICM4502P | OBICM4503P | OBICM4504P | OBICM4506P | OBICM4508P |

**• Regular**

| D TYPE G/H | | 2.0 | 3.0 | 4.0 | 6.0 | 8.0 |
|---|---|---|---|---|---|---|
| Ø 4.0 | Hex | OBICR4002P | OBICR4003P | OBICR4004P | OBICR4006P | OBICR4008P |
| Ø 4.5 | Hex | OBICR4502P | OBICR4503P | OBICR4504P | OBICR4506P | OBICR4508P |
| Ø 5.0 | Hex | OBICR5002P | OBICR5003P | OBICR5004P | OBICR5006P | OBICR5008P |
| Ø 6.0 | Hex | OBICR6002P | OBICR6003P | OBICR6004P | OBICR6006P | OBICR6008P |

## SM II Scanbody




| H(mm) | TYPE | CODE | 2.0 |
|---|---|---|---|
| 12 | Regular | OSBRP | Sub Merged |
| 16 | Mini | OSBMP | Sub Merged |

## Fixture Pick-up Impression Coping




- Packing Unit : Impression Coping Body + Guide Pin.
- Fixture Level Impression.
- Impression taken with open tray.
- Used when precise impression is required.
- Fastening by hand using 1.2 Hex Driver.

**Size Classification**

- Guide Pin : Regular(Purple), Mini(Non).
- Body : Regular(Non), Mini 3.5(Yellow), Min 3.0(Blue) Coloring.

• **Mini 3.0**

| D | TYPE \ L | 11 | 15 | 17 |
|---|---|---|---|---|
| Ø 4.0 | Hex | OSICPX**4011**HP | OSICPX**4015**HP | OSICPX**4017**HP |
| | Non-Hex | OSICPX**4011**NP | OSICPX**4015**NP | OSICPX**4017**NP |
| Ø 4.5 | Hex | OSICPX**4511**HP | OSICPX**4515**HP | OSICPX**4517**HP |
| | Non-Hex | OSICPX**4511**NP | OSICPX**4515**NP | OSICPX**4517**NP |

• **Mini 3.5**

| D | TYPE \ L | 11 | 15 | 17 |
|---|---|---|---|---|
| Ø 4.0 | Hex | OSICPM**4011**HP | OSICPM**4015**HP | OSICPM**4017**HP |
| | Non-Hex | OSICPM**4011**NP | OSICPM**4015**NP | OSICPM**4017**NP |
| Ø 4.5 | Hex | OSICPM**4511**HP | OSICPM**4515**HP | OSICPM**4517**HP |
| | Non-Hex | OSICPM**4511**NP | OSICPM**4515**NP | OSICPM**4517**NP |

• **Regular**

| D | TYPE \ L | 11 | 15 | 17 |
|---|---|---|---|---|
| Ø 4.0 | Hex | OSICP**4011**HP | OSICP**4015**HP | OSICP**4017**HP |
| | Non-Hex | OSICP**4011**NP | OSICP**4015**NP | OSICP**4017**NP |
| Ø 4.5 | Hex | OSICP**4511**HP | OSICP**4515**HP | OSICP**4517**HP |
| | Non-Hex | OSICP**4511**NP | OSICP**4515**NP | OSICP**4517**NP |
| Ø 5.0 | Hex | OSICP**5011**HP | OSICP**5015**HP | OSICP**5017**HP |
| | Non-Hex | OSICP**5011**NP | OSICP**5015**NP | OSICP**5017**NP |
| Ø 5.5 | Hex | OSICP**5511**HP | OSICP**5515**HP | OSICP**5517**HP |
| | Non-Hex | OSICP**5511**NP | OSICP**5515**NP | OSICP**5517**NP |
| Ø 6.0 | Hex | OSICP**6011**HP | OSICP**6015**HP | OSICP**6017**HP |
| | Non-Hex | OSICP**6011**NP | OSICP**6015**NP | OSICP**6017**NP |
| Ø 6.5 | Hex | OSICP**6511**HP | OSICP**6515**HP | OSICP**6517**HP |
| | Non-Hex | OSICP**6511**NP | OSICP**6515**NP | OSICP**6517**NP |
| Ø 7.0 | Hex | OSICP**7011**HP | OSICP**7015**HP | OSICP**7017**HP |
| | Non-Hex | OSICP**7011**NP | OSICP**7015**NP | OSICP**7017**NP |

## Fixture Transfer Impression Coping



- Packing Unit : Impression Coping Body + Guide Pin.
- Fixture Level impression.
- Impression taken with open tray.
- Used when precise impression is required.
- Fastening by hand using 1.2 Hex Driver.

**Size Classification**

- Guide Pin : Regular(Purple), Mini(None).
- Body : Regular(None), Mini 3.5(Yellow), Mini 3.0(Blue) Coloring.




**• Mini 3.0**

| D | TYPE\L | 11 | 14 | 16 |
|---|---|---|---|---|
| Ø 4.0 | Hex | OSICTX**4011HP** | OSICTX**4014HP** | OSICTX**4016HP** |
| | Non-Hex | OSICTX**4011NP** | OSICTX**4014NP** | OSICTX**4016NP** |
| Ø 4.5 | Hex | OSICTX**4511HP** | OSICTX**4514HP** | OSICTX**4516HP** |
| | Non-Hex | OSICTX**4511NP** | OSICTX**4514NP** | OSICTX**4516NP** |

**• Mini 3.5**

| D | TYPE\L | 11 | 14 | 16 |
|---|---|---|---|---|
| Ø 4.0 | Hex | OSICTM**4011HP** | OSICTM**4014HP** | OSICTM**4016HP** |
| | Non-Hex | OSICTM**4011NP** | OSICTM**4014NP** | OSICTM**4016NP** |
| Ø 4.5 | Hex | OSICTM**4511HP** | OSICTM**4514HP** | OSICTM**4516HP** |
| | Non-Hex | OSICTM**4511NP** | OSICTM**4514NP** | OSICTM**4516NP** |

**• Regular**

| D | TYPE\L | 11 | 14 | 16 |
|---|---|---|---|---|
| Ø 4.0 | Hex | OSICT**4011HP** | OSICT**4015HP** | OSICT**4017HP** |
| | Non-Hex | OSICT**4011NP** | OSICT**4015NP** | OSICT**4017NP** |
| Ø 4.5 | Hex | OSICT**4511HP** | OSICT**4515HP** | OSICT**4517HP** |
| | Non-Hex | OSICT**4511NP** | OSICT**4515NP** | OSICT**4517NP** |
| Ø 5.0 | Hex | OSICT**5011HP** | OSICT**5015HP** | OSICT**5017HP** |
| | Non-Hex | OSICT**5011NP** | OSICT**5015NP** | OSICT**5017NP** |
| Ø 5.5 | Hex | OSICT**5511HP** | OSICT**5515HP** | OSICT**5517HP** |
| | Non-Hex | OSICT**5511NP** | OSICT**5515NP** | OSICT**5517NP** |
| Ø 6.0 | Hex | OSICT**6011HP** | OSICT**6015HP** | OSICT**6017HP** |
| | Non-Hex | OSICT**6011NP** | OSICT**6015NP** | OSICT**6017NP** |
| Ø 6.5 | Hex | OSICT**6511HP** | OSICT**6515HP** | OSICT**6517HP** |
| | Non-Hex | OSICT**6511NP** | OSICT**6515NP** | OSICT**6517NP** |
| Ø 7.0 | Hex | OSICT**7011HP** | OSICT**7015HP** | OSICT**7017HP** |
| | Non-Hex | OSICT**7011NP** | OSICT**7015NP** | OSICT**7017NP** |

## Fixture Lab Analog




- Packing Unit : Fixture Analog.
- Provide an anchor point for fabricating implant prosthetics when operating on a working model.
- 1.2 Hex Driver / with 10 N of torque.

**Size Classification**

- Body : Mini 3.0(Blue), Mini 3.5(Yellow), Regular 4.0(Purple) Coloring.

**• Mini 3.0**

| D | CODE |
|---|---|
| Ø 3.0 | OSTLA**300** |

**• Mini 3.5**

| D | CODE |
|---|---|
| Ø 3.5 | OSTLA**350** |

**• Regular 4.0**

| D | CODE |
|---|---|
| Ø 4.0 | OSTLA**400** |


## Ball Lab Analog



- Lab analog for Ball Abutment
- Ball Abutment Common Use.

| D | CODE |
|---|---|
| Ø 3.5 | OSBLA400 |

## Cemented Abutment




- Packing Unit : Abutment + Abutment Screw.
- Loaded with 1.2 Hex Driver.
- Recommended torque: 20Ncm (Mini), 30Ncm (Regular)
- Titanium Screw Catalog Code : OAS1610(Mini), OAS2008(Regular)
- Re-tighten with a torque of 30Ncm after 10 minutes of initial placement.

### • Mini

| D | L | TYPE\G/H | 1.0 | 2.0 | 3.0 | 4.0 | 5.0 | 6.0 | 7.0 |
|---|---|---|---|---|---|---|---|---|---|
| Ø 4.0 | 4.0 | Hex | OSCAM4041HTP | OSCAM4042HTP | OSCAM4043HTP | OSCAM4044HTP | OSCAM4045HTP | OSCAM4046HTP | OSCAM4047HTP |
| | | Non-Hex | OSCAM4041NTP | OSCAM4042NTP | OSCAM4043NTP | OSCAM4044NTP | OSCAM4045NTP | OSCAM4046NTP | OSCAM4047NTP |
| | 5.5 | Hex | OSCAM4051HTP | OSCAM4052HTP | OSCAM4053HTP | OSCAM4054HTP | OSCAM4055HTP | OSCAM4056HTP | OSCAM4057HTP |
| | | Non-Hex | OSCAM4051NTP | OSCAM4052NTP | OSCAM4053NTP | OSCAM4054NTP | OSCAM4055NTP | OSCAM4056NTP | OSCAM4057NTP |
| | 7.0 | Hex | OSCAM4071HTP | OSCAM4072HTP | OSCAM4073HTP | OSCAM4074HTP | OSCAM4075HTP | OSCAM4076HTP | OSCAM4077HTP |
| | | Non-Hex | OSCAM4071NTP | OSCAM4072NTP | OSCAM4073NTP | OSCAM4074NTP | OSCAM4075NTP | OSCAM4076NTP | OSCAM4077NTP |
| Ø 4.5 | 4.0 | Hex | OSCAM4541HTP | OSCAM4542HTP | OSCAM4543HTP | OSCAM4544HTP | OSCAM4545HTP | OSCAM4546HTP | OSCAM4547HTP |
| | | Non-Hex | OSCAM4541NTP | OSCAM4542NTP | OSCAM4543NTP | OSCAM4544NTP | OSCAM4545NTP | OSCAM4546NTP | OSCAM4547NTP |
| | 5.5 | Hex | OSCAM4551HTP | OSCAM4552HTP | OSCAM4553HTP | OSCAM4554HTP | OSCAM4555HTP | OSCAM4556HTP | OSCAM4557HTP |
| | | Non-Hex | OSCAM4551NTP | OSCAM4552NTP | OSCAM4553NTP | OSCAM4554NTP | OSCAM4555NTP | OSCAM4556NTP | OSCAM4557NTP |
| | 7.0 | Hex | OSCAM4571HTP | OSCAM4572HTP | OSCAM4573HTP | OSCAM4574HTP | OSCAM4575HTP | OSCAM4576HTP | OSCAM4577HTP |
| | | Non-Hex | OSCAM4571NTP | OSCAM4572NTP | OSCAM4573NTP | OSCAM4574NTP | OSCAM4575NTP | OSCAM4576NTP | OSCAM4577NTP |
| Ø 5.0 | 4.0 | Hex | OSCAM5041HTP | OSCAM5042HTP | OSCAM5043HTP | OSCAM5044HTP | OSCAM5045HTP | OSCAM5046HTP | OSCAM5047HTP |
| | | Non-Hex | OSCAM5041NTP | OSCAM5042NTP | OSCAM5043NTP | OSCAM5044NTP | OSCAM5045NTP | OSCAM5046NTP | OSCAM5047NTP |
| | 5.5 | Hex | OSCAM5051HTP | OSCAM5052HTP | OSCAM5053HTP | OSCAM5054HTP | OSCAM5055HTP | OSCAM5056HTP | OSCAM5057HTP |
| | | Non-Hex | OSCAM5051NTP | OSCAM5052NTP | OSCAM5053NTP | OSCAM4054NTP | OSCAM5055NTP | OSCAM5056NTP | OSCAM5057NTP |
| | 7.0 | Hex | OSCAM5071HTP | OSCAM5072HTP | OSCAM5073HTP | OSCAM5074HTP | OSCAM5075HTP | OSCAM5076HTP | OSCAM5077HTP |
| | | Non-Hex | OSCAM5071NTP | OSCAM5072NTP | OSCAM5073NTP | OSCAM5074NTP | OSCAM5075NTP | OSCAM5076NTP | OSCAM5077NTP |


### • Regular

| D | L | TYPE\G/H | 1.0 | 2.0 | 3.0 | 4.0 | 5.0 | 6.0 | 7.0 |
|---|---|---|---|---|---|---|---|---|---|
| Ø 4.0 | 4.0 | Hex | OSCA4041HTP | OSCA4042HTP | OSCA4043HTP | OSCA4044HTP | OSCA4045HTP | OSCA4046HTP | OSCA4047HTP |
| | | Non-Hex | OSCA4041NTP | OSCA4042NTP | OSCA4043NTP | OSCA4044NTP | OSCA4045NTP | OSCA4046NTP | OSCA4047NTP |
| | 5.5 | Hex | OSCA4051HTP | OSCA4052HTP | OSCA4053HTP | OSCA4054HTP | OSCA4055HTP | OSCA4056HTP | OSCA4057HTP |
| | | Non-Hex | OSCA4051NTP | OSCA4052NTP | OSCA4053NTP | OSCA4054NTP | OSCA4055NTP | OSCA4056NTP | OSCA4057NTP |
| | 7.0 | Hex | OSCA4071HTP | OSCA4072HTP | OSCA4073HTP | OSCA4074HTP | OSCA4075HTP | OSCA4076HTP | OSCA4077HTP |
| | | Non-Hex | OSCA4071NTP | OSCA4072NTP | OSCA4073NTP | OSCA4074NTP | OSCA4075NTP | OSCA4076NTP | OSCA4077NTP |
| Ø 4.5 | 4.0 | Hex | OSCA4541HTP | OSCA4542HTP | OSCA4543HTP | OSCA4544HTP | OSCA4545HTP | OSCA4546HTP | OSCA4547HTP |
| | | Non-Hex | OSCA4541NTP | OSCA4542NTP | OSCA4543NTP | OSCA4544NTP | OSCA4545NTP | OSCA4546NTP | OSCA4547NTP |
| | 5.5 | Hex | OSCA4551HTP | OSCA4552HTP | OSCA4553HTP | OSCA4554HTP | OSCA4555HTP | OSCA4556HTP | OSCA4557HTP |
| | | Non-Hex | OSCA4551NTP | OSCA4552NTP | OSCA4553NTP | OSCA4554NTP | OSCA4555NTP | OSCA4556NTP | OSCA4557NTP |
| | 7.0 | Hex | OSCA4571HTP | OSCA4572HTP | OSCA4573HTP | OSCA4574HTP | OSCA4575HTP | OSCA4576HTP | OSCA4577HTP |
| | | Non-Hex | OSCA4571NTP | OSCA4572NTP | OSCA4573NTP | OSCA4574NTP | OSCA4575NTP | OSCA4576NTP | OSCA4577NTP |
| Ø 5.0 | 4.0 | Hex | OSCA5041HTP | OSCA5042HTP | OSCA5043HTP | OSCA5044HTP | OSCA5045HTP | OSCA5046HTP | OSCA5047HTP |
| | | Non-Hex | OSCA5041NTP | OSCA5042NTP | OSCA5043NTP | OSCA5044NTP | OSCA5045NTP | OSCA5046NTP | OSCA5047NTP |
| | 5.5 | Hex | OSCA5051HTP | OSCA5052HTP | OSCA5053HTP | OSCA5054HTP | OSCA5055HTP | OSCA5056HTP | OSCA5057HTP |
| | | Non-Hex | OSCA5051NTP | OSCA5052NTP | OSCA5053NTP | OSCA4054NTP | OSCA5055NTP | OSCA5056NTP | OSCA5057NTP |
| | 7.0 | Hex | OSCA5071HTP | OSCA5072HTP | OSCA5073HTP | OSCA5074HTP | OSCA5075HTP | OSCA5076HTP | OSCA5077HTP |
| | | Non-Hex | OSCA5071NTP | OSCA5072NTP | OSCA5073NTP | OSCA5074NTP | OSCA5075NTP | OSCA5076NTP | OSCA5077NTP |

• **Regular**

| D | L | TYPE\G/H | 1.0 | 2.0 | 3.0 | 4.0 | 5.0 | 6.0 | 7.0 |
|---|---|---|---|---|---|---|---|---|---|
| Ø 5.0 | 4.0 | Hex | OSCA5541HTP | OSCA5542HTP | OSCA5543HTP | OSCA5544HTP | OSCA5545HTP | OSCA5546HTP | OSCA5547HTP |
| | | Non-Hex | OSCA5541NTP | OSCA5542NTP | OSCA5543NTP | OSCA5544NTP | OSCA5545NTP | OSCA5546NTP | OSCA5547NTP |
| | 5.5 | Hex | OSCA5551HTP | OSCA5552HTP | OSCA5553HTP | OSCA5554HTP | OSCA5555HTP | OSCA5556HTP | OSCA5557HTP |
| | | Non-Hex | OSCA5551NTP | OSCA5552NTP | OSCA5553NTP | OSCA5554NTP | OSCA5555NTP | OSCA5556NTP | OSCA5557NTP |
| | 7.0 | Hex | OSCA5571HTP | OSCA5572HTP | OSCA5573HTP | OSCA5574HTP | OSCA5575HTP | OSCA5576HTP | OSCA5577HTP |
| | | Non-Hex | OSCA5571NTP | OSCA5572NTP | OSCA5573NTP | OSCA5574NTP | OSCA5575NTP | OSCA5576NTP | OSCA5577NTP |
| Ø 6.0 | 4.0 | Hex | OSCA6041HTP | OSCA6042HTP | OSCA6043HTP | OSCA6044HTP | OSCA6045HTP | OSCA6046HTP | OSCA6047HTP |
| | | Non-Hex | OSCA6041NTP | OSCA6042NTP | OSCA6043NTP | OSCA6044NTP | OSCA6045NTP | OSCA6046NTP | OSCA6047NTP |
| | 5.5 | Hex | OSCA6051HTP | OSCA6052HTP | OSCA6053HTP | OSCA6054HTP | OSCA6055HTP | OSCA6056HTP | OSCA6057HTP |
| | | Non-Hex | OSCA6051NTP | OSCA6052NTP | OSCA6053NTP | OSCA6054NTP | OSCA6055NTP | OSCA6056NTP | OSCA6057NTP |
| | 7.0 | Hex | OSCA6071HTP | OSCA6072HTP | OSCA6073HTP | OSCA6074HTP | OSCA6075HTP | OSCA6076HTP | OSCA6077HTP |
| | | Non-Hex | OSCA6071NTP | OSCA6072NTP | OSCA6073NTP | OSCA6074NTP | OSCA6075NTP | OSCA6076NTP | OSCA6077NTP |
| Ø 6.5 | 4.0 | Hex | OSCA6541HTP | OSCA6542HTP | OSCA6543HTP | OSCA6544HTP | OSCA6545HTP | OSCA6546HTP | OSCA6547HTP |
| | | Non-Hex | OSCA6541NTP | OSCA6542NTP | OSCA6543NTP | OSCA6544NTP | OSCA6545NTP | OSCA6546NTP | OSCA6547NTP |
| | 5.5 | Hex | OSCA6551HTP | OSCA6552HTP | OSCA6553HTP | OSCA6554HTP | OSCA6555HTP | OSCA6556HTP | OSCA6557HTP |
| | | Non-Hex | OSCA6551NTP | OSCA6552NTP | OSCA6553NTP | OSCA6554NTP | OSCA6555NTP | OSCA6556NTP | OSCA6557NTP |
| | 7.0 | Hex | OSCA6571HTP | OSCA6572HTP | OSCA6573HTP | OSCA6574HTP | OSCA6575HTP | OSCA6576HTP | OSCA6577HTP |
| | | Non-Hex | OSCA6571NTP | OSCA6572NTP | OSCA6573NTP | OSCA6574NTP | OSCA6575NTP | OSCA6576NTP | OSCA6577NTP |
| Ø 7.0 | 4.0 | Hex | OSCA7041HTP | OSCA7042HTP | OSCA7043HTP | OSCA7044HTP | OSCA7045HTP | OSCA7046HTP | OSCA7047HTP |
| | | Non-Hex | OSCA7041NTP | OSCA7042NTP | OSCA7043NTP | OSCA7044NTP | OSCA7045NTP | OSCA7046NTP | OSCA7047NTP |
| | 5.5 | Hex | OSCA7051HTP | OSCA7052HTP | OSCA7053HTP | OSCA7054HTP | OSCA7055HTP | OSCA7056HTP | OSCA7057HTP |
| | | Non-Hex | OSCA7051NTP | OSCA7052NTP | OSCA7053NTP | OSCA7054NTP | OSCA7055NTP | OSCA7056NTP | OSCA7057NTP |
| | 7.0 | Hex | OSCA7071HTP | OSCA7072HTP | OSCA7073HTP | OSCA7074HTP | OSCA7075HTP | OSCA7076HTP | OSCA7077HTP |
| | | Non-Hex | OSCA7071NTP | OSCA7072NTP | OSCA7073NTP | OSCA7074NTP | OSCA7075NTP | OSCA7076NTP | OSCA7077NTP |

## Angled Abutment




- Packing Unit : Abutment + Abutment Screw.
- Suitable for up to 23°Angle.
- A, B Type and Non-Hex, selected according to the Fixture Hex direction.
- Loaded with 1.2 Hex Driver.
- Recommended torque: 20Ncm (Mini), 30Ncm (Regular).
- Titanium Screw Catalog Code : OAS1610(Mini), OAS2008(Regular).
- Re-tighten with a torque of 30Ncm after 10 minutes of initial placement.


• **Mini**

| D | TYPE\G/H | 2.0 | 3.0 | 4.0 |
|---|---|---|---|---|
| Ø 4.0 | Hex(A Type) | OSAAM4002ATP | OSAAM4003ATP | OSAAM4004ATP |
| | Hex(B Type) | OSAAM4002BTP | OSAAM4003BTP | OSAAM4004BTP |
| | Non-Hex | OSAAM4002NTP | OSAAM4003NTP | OSAAM4004NTP |
| Ø 4.5 | Hex(A Type) | OSAAM4502ATP | OSAAM4503ATP | OSAAM4504ATP |
| | Hex(B Type) | OSAAM4502BTP | OSAAM4503BTP | OSAAM4504BTP |
| | Non-Hex | OSAAM4502NTP | OSAAM4503NTP | OSAAM4504NTP |

• **Regular**

| D | TYPE\G/H | 2.0 | 3.0 | 4.0 |
|---|---|---|---|---|
| Ø 4.5 | Hex(A Type) | OSAA4502ATP | OSAA4503ATP | OSAA4504ATP |
| | Hex(B Type) | OSAA4502BTP | OSAA4503BTP | OSAA4504BTP |
| | Non-Hex | OSAA4502NTP | OSAA4503NTP | OSAA4504NTP |
| Ø 5.0 | Hex(A Type) | OSAA5002ATP | OSAA5003ATP | OSAA5004ATP |
| | Hex(B Type) | OSAA5002BTP | OSAA5003BTP | OSAA5004BTP |
| | Non-Hex | OSAA5002NTP | OSAA5003NTP | OSAA5004NTP |
| Ø 6.0 | Hex(A Type) | OSAA6002ATP | OSAA6003ATP | OSAA6004ATP |
| | Hex(B Type) | OSAA6002BTP | OSAA6003BTP | OSAA6004BTP |
| | Non-Hex | OSAA6002NTP | OSAA6003NTP | OSAA6004NTP |

## Milling Abutment




- Packing Unit: Abutment + Abutment Screw
- Use a 1.2 Hex Driver.
- Recommended tightening torque: 20Ncm (Mini), 30Ncm (Regular).
- Titanium screw: OAS1610 (Mini), OAS2008 (Regular).
- Re-tighten with a torque of 30Ncm after 10 minutes of initial placement.


**• Mini**

| D | TYPE\G/H | 1.0 | 2.0 | 3.0 | 4.0 | 5.0 | 6.0 | 7.0 |
|---|---|---|---|---|---|---|---|---|
| Ø 4.0 | Hex | OSMAM4010HTP | OSMAM4020HTP | OSMAM4030HTP | OSMAM4040HTP | OSMAM4050HTP | OSMAM4060HTP | OSMAM4070HTP |
| | Non-Hex | OSMAM4010NTP | OSMAM4020NTP | OSMAM4030NTP | OSMAM4040NTP | OSMAM4050NTP | OSMAM4060NTP | OSMAM4070NTP |
| Ø 4.5 | Hex | OSMAM4510HTP | OSMAM4520HTP | OSMAM4530HTP | OSMAM4540HTP | OSMAM4550HTP | OSMAM4560HTP | OSMAM4570HTP |
| | Non-Hex | OSMAM4510NTP | OSMAM4520NTP | OSMAM4530NTP | OSMAM4540NTP | OSMAM4550NTP | OSMAM4560NTP | OSMAM4570NTP |

**• Regular**

| D | TYPE\G/H | 1.0 | 2.0 | 3.0 | 4.0 | 5.0 | 6.0 | 7.0 |
|---|---|---|---|---|---|---|---|---|
| Ø 4.5 | Hex | OSMA4510HTP | OSMA4520HTP | OSMA4530HTP | OSMA4540HTP | OSMA4550HTP | OSMA4560HTP | OSMA4570HTP |
| | Non-Hex | OSMA4510NTP | OSMA4520NTP | OSMA4530NTP | OSMA4540NTP | OSMA4550NTP | OSMA4560NTP | OSMA4570NTP |
| Ø 5.0 | Hex | OSMA5010HTP | OSMA5020HTP | OSMA5030HTP | OSMA5040HTP | OSMA5050HTP | OSMA5060HTP | OSMA5070HTP |
| | Non-Hex | OSMA5010NTP | OSMA5020NTP | OSMA5030NTP | OSMA5040NTP | OSMA5050NTP | OSMA5060NTP | OSMA5070NTP |
| Ø 5.5 | Hex | OSMA5510HTP | OSMA5520HTP | OSMA5530HTP | OSMA5540HTP | OSMA5550HTP | OSMA5560HTP | OSMA5570HTP |
| | Non-Hex | OSMA5510NTP | OSMA5520NTP | OSMA5530NTP | OSMA5540NTP | OSMA5550NTP | OSMA5560NTP | OSMA5570NTP |
| Ø 6.0 | Hex | OSMA6010HTP | OSMA6020HTP | OSMA6030HTP | OSMA6040HTP | OSMA6050HTP | OSMA6060HTP | OSMA6070HTP |
| | Non-Hex | OSMA6010NTP | OSMA6020NTP | OSMA6030NTP | OSMA6040NTP | OSMA6050NTP | OSMA6060NTP | OSMA6070NTP |
| Ø 6.5 | Hex | OSMA6510HTP | OSMA6520HTP | OSMA6530HTP | OSMA6540HTP | OSMA6550HTP | OSMA6560HTP | OSMA6570HTP |
| | Non-Hex | OSMA6510NTP | OSMA6520NTP | OSMA6530NTP | OSMA6540NTP | OSMA6550NTP | OSMA6560NTP | OSMA6570NTP |
| Ø 7.0 | Hex | OSMA7010HTP | OSMA7020HTP | OSMA7030HTP | OSMA7040HTP | OSMA7050HTP | OSMA7060HTP | OSMA7070HTP |
| | Non-Hex | OSMA7010NTP | OSMA7020NTP | OSMA7030NTP | OSMA7040NTP | OSMA7050NTP | OSMA7060NTP | OSMA7070NTP |

## Temporary Abutment




- Packing Unit : Temporary Abutment + Abutment Screw.
- Form and use to make temporary prostheses.
- Loaded with 1.2 Hex Driver.
- Recommended torque: 20Ncm (Mini), 30Ncm (Regular).
- Titanium Screw Catalog Code : OAS1610(Mini), OAS2008(Regular).
- Re-tighten with a torque of 30Ncm after 10 minutes of initial placement.

**• Mini**

| D | TYPE \ G/H | 1.0 | 2.0 | 3.0 | 4.0 |
|---|---|---|---|---|---|
| Ø 4.0 | Hex | OSTAM4010HP | OSTAM4020HP | OSTAM4030HP | OSTAM4040HP |
| | Non-Hex | OSTAM4010NP | OSTAM4020NP | OSTAM4030NP | OSTAM4040NP |

**• Regular**

| D | TYPE \ G/H | 1.0 | 2.0 | 3.0 | 4.0 |
|---|---|---|---|---|---|
| Ø 4.0 | Hex | OSTA4010HP | OSTA4020HP | OSTA4030HP | OSTA4040HP |
| | Non-Hex | OSTA4010NP | OSTA4020NP | OSTA4030NP | OSTA4040NP |
| Ø 4.5 | Hex | OSTA4510HP | OSTA4520HP | OSTA4530HP | OSTA4540HP |
| | Non-Hex | OSTA4510NP | OSTA4520NP | OSTA4530NP | OSTA4540NP |
| Ø 5.0 | Hex | OSTA5010HP | OSTA5020HP | OSTA5030HP | OSTA5040HP |
| | Non-Hex | OSTA5010NP | OSTA5020NP | OSTA5030NP | OSTA5040NP |

## CCM Casting Abutment




- Packing Unit: Casting Abutment + Abutment Screw.
- Used when casting with non-precious metal alloy (Cr-Co-Mo) and fabricating a custom prosthesis.
- Use a 1.2 Hex Driver.
- Recommended tightening torque: 20Ncm (Mini), 30Ncm (Regular).
- Titanium screw: OAS1610 (Mini), OAS2008 (Regular)
- Re-tighten with a torque of 30Ncm after 10 minutes of initial placement.


**· Mini**

| D | TYPE\G/H | 1.0 | 2.0 | 3.0 | 4.0 |
|---|---|---|---|---|---|
| Ø 3.5 | Hex | OSCA**3501**HP | OSCA**3502**HP | OSCA**3503**HP | OSCA**3504**HP |
| | Non-Hex | OSCA**3501**NP | OSCA**3502**NP | OSCA**3503**NP | OSCA**3504**NP |

**· Regular**

| D | TYPE\G/H | 1.0 | 2.0 | 3.0 | 4.0 |
|---|---|---|---|---|---|
| Ø 4.5 | Hex | OSCA**4501**HP | OSCA**4502**HP | OSCA**4503**HP | OSCA**4504**HP |
| | Non-Hex | OSCA**4501**NP | OSCA**4502**NP | OSCA**4503**NP | OSCA**4504**NP |
| Ø 5.0 | Hex | OSCA**5001**HP | OSCA**5002**HP | OSCA**5003**HP | OSCA**5004**HP |
| | Non-Hex | OSCA**5001**NP | OSCA**5002**NP | OSCA**5003**NP | OSCA**5004**NP |

## Solid Abutment




- Packing Unit : Abutment + Protect Cap.
- Loaded with 1.2 Hex Driver.
- Recommended torque: 30Ncm(Mini & Regular).
- Surface : SLA Surface + UV. (Effect of maximizing hydrophilicity by expanding the UV irradiation absorption area)
- Outer Driver : Short(ORDS), Long(ORDL).
- Re-tighten with a torque of 30Ncm after 10 minutes of initial placement.

**· Mini**

| D | G/H＼L | 4.0 | 5.5 | 7.0 |
|---|---|---|---|---|
| Ø 4.0 | 1.0 | OSSAM**4041**TP | OSSAM**4051**TP | OSSAM**4071**TP |
| | 2.0 | OSSAM**4042**TP | OSSAM**4052**TP | OSSAM**4072**TP |
| | 3.0 | OSSAM**4043**TP | OSSAM**4053**TP | OSSAM**4073**TP |
| | 4.0 | OSSAM**4044**TP | OSSAM**4054**TP | OSSAM**4074**TP |
| | 5.0 | OSSAM**4045**TP | OSSAM**4055**TP | OSSAM**4075**TP |
| Ø 4.5 | 1.0 | OSSAM**4541**TP | OSSAM**4551**TP | OSSAM**4571**TP |
| | 2.0 | OSSAM**4542**TP | OSSAM**4552**TP | OSSAM**4572**TP |
| | 3.0 | OSSAM**4543**TP | OSSAM**4553**TP | OSSAM**4573**TP |
| | 4.0 | OSSAM**4544**TP | OSSAM**4554**TP | OSSAM**4574**TP |
| | 5.0 | OSSAM**4545**TP | OSSAM**4555**TP | OSSAM**4575**TP |

**· Regular**

| D | G/H＼L | 4.0 | 5.5 | 7.0 |
|---|---|---|---|---|
| Ø 4.0 | 1.0 | OSSA**4041**TP | OSSA**4051**TP | OSSA**4071**TP |
| | 2.0 | OSSA**4042**TP | OSSA**4052**TP | OSSA**4072**TP |
| | 3.0 | OSSA**4043**TP | OSSA**4053**TP | OSSA**4073**TP |
| | 4.0 | OSSA**4044**TP | OSSA**4054**TP | OSSA**4074**TP |
| | 5.0 | OSSA**4045**TP | OSSA**4055**TP | OSSA**4075**TP |
| Ø 4.5 | 1.0 | OSSA**4541**TP | OSSA**4551**TP | OSSA**4571**TP |
| | 2.0 | OSSA**4542**TP | OSSA**4552**TP | OSSA**4572**TP |
| | 3.0 | OSSA**4543**TP | OSSA**4553**TP | OSSA**4573**TP |
| | 4.0 | OSSA**4544**TP | OSSA**4554**TP | OSSA**4574**TP |
| | 5.0 | OSSA**4545**TP | OSSA**4555**TP | OSSA**4575**TP |
| Ø 5.0 | 1.0 | OSSA**5041**TP | OSSA**5051**TP | OSSA**5071**TP |
| | 2.0 | OSSA**5042**TP | OSSA**5052**TP | OSSA**5072**TP |
| | 3.0 | OSSA**5043**TP | OSSA**5053**TP | OSSA**5073**TP |
| | 4.0 | OSSA**5044**TP | OSSA**5054**TP | OSSA**5074**TP |
| | 5.0 | OSSA**5045**TP | OSSA**5055**TP | OSSA**5075**TP |
| Ø 6.0 | 1.0 | OSSA**6041**TP | OSSA**6051**TP | OSSA**6071**TP |
| | 2.0 | OSSA**6042**TP | OSSA**6052**TP | OSSA**6072**TP |
| | 3.0 | OSSA**6043**TP | OSSA**6053**TP | OSSA**6073**TP |
| | 4.0 | OSSA**6044**TP | OSSA**6054**TP | OSSA**6074**TP |
| | 5.0 | OSSA**6045**TP | OSSA**6055**TP | OSSA**6075**TP |
| Ø 7.0 | 1.0 | OSSA**7041**TP | OSSA**7051**TP | OSSA**7071**TP |
| | 2.0 | OSSA**7042**TP | OSSA**7052**TP | OSSA**7072**TP |
| | 3.0 | OSSA**7043**TP | OSSA**7053**TP | OSSA**7073**TP |
| | 4.0 | OSSA**7044**TP | OSSA**7054**TP | OSSA**7074**TP |
| | 5.0 | OSSA**7045**TP | OSSA**7055**TP | OSSA**7075**TP |

## Solid Outer Driver


- Outer Driver used for Ø4.0
- If a significant portion of the Ø 4.5 abutment is removed, the hex driver socket may become unusable.



- Mini

| SIZE | D | CODE |
|---|---|---|
| S | Ø 4.0 | ORDS4000 |
| L | | ORDL4000 |

## Loc Abutment




- Compatible with Zest's Locator Abutment for TS.
- Suitable for up to 40° placement angles.
- 1.5mm low vertical diameter, various and stable attachment configuration.
- Recommended torque: 30Ncm(Mini & Regular).
- Used for Denture Lock in male parts.
- Used in accordance with 3.0 and 3.5 Fixtures Mini Connection.
- Re-tighten with a torque of 30Ncm after 10 minutes of initial placement.

- Mini 3.0

| D\G/H | 1.0 | 2.0 | 3.0 | 4.0 | 5.0 |
|---|---|---|---|---|---|
| Ø 3.7 | OLAX4001T | OLAX4002T | OLAX4003T | OLAX4004T | OLAX4005T |


- Mini 3.5

| D\G/H | 1.0 | 2.0 | 3.0 | 4.0 | 5.0 |
|---|---|---|---|---|---|
| Ø 3.7 | OLAM4001T | OLAM4002T | OLAM4003T | OLAM4004T | OLAM4005T |

- Regular

| D\G/H | 1.0 | 2.0 | 3.0 | 4.0 | 5.0 |
|---|---|---|---|---|---|
| Ø 3.7 | OLA4001T | OLA4002T | OLA4003T | OLA4004T | OLA4005T |

## Ball Abutment




- An abutment for overdentures using an O-ring Attachment.
- Angulation compensation up to 20°. Tightened using a dedicated Outer Driver.
- Recommended tightening torque: 30Ncm (min)/regular).
- For Mini Connection, order to match the 3.0 and 3.5 Fixtures.
- Re-tighten with a torque of 30Ncm after 10 minutes of initial placement.

**· Mini 3.0**

| D\ G/H | 1.0 | 2.0 | 3.0 | 4.0 | 5.0 |
|---|---|---|---|---|---|
| Ø 3.5 | OBAX4001T | OBAX4002T | OBAX4003T | OBAX4004T | OBAX4005T |

**· Mini 3.5**

| D\ G/H | 1.0 | 2.0 | 3.0 | 4.0 | 5.0 |
|---|---|---|---|---|---|
| Ø 3.7 | OBAM4001T | OBAM4002T | OBAM4003T | OBAM4004T | OBAM4005T |

**· Regular**

| D\ G/H | 1.0 | 2.0 | 3.0 | 4.0 | 5.0 |
|---|---|---|---|---|---|
| Ø 3.7 | OBA4001T | OBA4002T | OBA4003T | OBA4004T | OBA4005T |

## Ball Abutment Driver




- Outer Driver for Ball Abutment.
- Used for Carrying and Fastening of Ball Abutment.
- Ball Abutment Common Use.

| D | CODE |
|---|---|
| S | OBDT2400S |
| L | OBDT2400L |

## Solid Protect Cap



- Solid abutment protection.
- Can be used as a temporary crown base.

| D \ L | 4.0 | 5.5 | 7.0 |
|---|---|---|---|
| Ø 4.0 | OSSPC404 | OSSPC405 | OSSPC407 |
| Ø 4.5 | OSSPC454 | OSSPC455 | OSSPC457 |
| Ø 5.0 | OSSPC504 | OSSPC505 | OSSPC507 |
| Ø 6.0 | OSSPC604 | OSSPC605 | OSSPC607 |
| Ø 7.0 | OSSPC704 | OSSPC705 | OSSPC707 |

# TN

TN UV Deep Thread Blade Type Fixture

# TN UV Deep Thread Blade Type Fixture



- Submerged-type implant featuring an Internal Hex and 11° Tapered Connection for enhanced structural stability.
- Optimized thread geometry designed for efficient cutting and stable engagement.
- Tapered body combined with advanced thread design delivers excellent primary stability.
- Deep blade-thread structure provides strong self-threading capability.
- Ideal for soft bone and sinus cases, offering predictable clinical outcomes.
- Micro-groove design increases the functional surface area to support superior osseointegration.
- High primary stability suitable for immediate loading, even in soft bone conditions.
- Surface: SLA Surface + UV
  (The structure expands the UV absorption area, maximizing hydrophilicity.)
- Recommended insertion torque: ≥ 40 Ncm.
- For posterior single-tooth restorations, Ø4.5 mm or wider fixtures are recommended.



- **Mini**

| D \ L | 7 | 8.5 | 10 | 11.5 | 13 | 16 |
|---|---|---|---|---|---|---|
| Ø 3.0 | TBF3007QNP | TBF3008QNP | TBF3010QNP | TBF3011QNP | TBF3013QNP | TBF3016QNP |
| Ø 3.5 | TBF3507QNP | TBF3508QNP | TBF3510QNP | TBF3511QNP | TBF3513QNP | TBF3516QNP |



- **Regular**

| D \ L | 7 | 8.5 | 10 | 11.5 | 13 | 16 |
|---|---|---|---|---|---|---|
| Ø 4.0 | TBF4007QNP | TBF4008QNP | TBF4010QNP | TBF4011QNP | TBF4013QNP | TBF4016QNP |
| Ø 4.5 | TBF4507QNP | TBF4508QNP | TBF4510QNP | TBF4511QNP | TBF4513QNP | TBF4516QNP |
| Ø 5.0 | TBF5007QNP | TBF5008QNP | TBF5010QNP | TBF5011QNP | TBF5013QNP | TBF5016QNP |
| Ø 5.5 | TBF5507QNP | TBF5508QNP | TBF5510QNP | TBF5511QNP | TBF5513QNP | TBF5516QNP |



- **Ultra-Wide**


| D \ L | 7 | 8.5 | 10 | 11.5 | 13 | 16 |
|---|---|---|---|---|---|---|
| Ø 6.0 | TBF6007QNP | TBF6008QNP | TBF6010QNP | TBF6011QNP | TBF6013QNP | TBF6016QNP |
| Ø 6.5 | TBF6507QNP | TBF6508QNP | TBF6510QNP | TBF6511QNP | TBF6513QNP | TBF6516QNP |

※ All prosthetic components and surgical kits designed for the SM II series are fully compatible with the TN series.

# SM I

**SM I UV Fixture (No-Mount)**
**SM I SLA Fixture (No-Mount)**
Cover Screw
Healing Abutment
SM I Scanbody
Fixture Pick-up Impression Coping
Fixture Transfer Impression Coping
Solid Lab Analog
Excellent Solid Lab Analog
Cemented Abutment
Excellent Cemented Abutment
Angled Abutment
Milling Abutment
Excellent Temporary Abutment
CCM Casting Abutment
Solid Abutment
Excellent Solid Abutment
Fixture Pin
Excellent Solid Abutment

# SM I Flowchart



# SM I UV Fixture (No-Mount)



- An implant with an Internal Octagon and 8° Taper Connection structure.
- SLA surface treatment: excellent for bone adhesion.
- A body design that is easy to adjust during insertion and has excellent primary stability.
- Corkscrew Thread for a powerful self-threading effect.
- Has Korea's only square-shaped Body Thread, effective for dispersing stress from chewing pressure.
- Increases convenience as it allows for one-stage surgery instead of two-stage surgery.
- Surface: SLA Surface + UV.
  (By increasing the UV molecular absorption area, the hydrophilic effect is maximized.)
- Recommended insertion torque: 40Ncm or less.
- For single molar cases, use a fixture of Ø 4.5mm or larger.

## UV Fixture



**• Regular ( G / H : 1.8 )**

| D \ L | 7 | 8.5 | 10 | 11.5 | 13 |
|---|---|---|---|---|---|
| Ø 3.5 | - | ITF350818QNP | ITF351018QNP | ITF351118QNP | ITF351318QNP |
| Ø 4.0 | ITF400718QNP | ITF400818QNP | ITF401018QNP | ITF401118QNP | ITF401318QNP |
| Ø 4.5 | ITF450718QNP | ITF450818QNP | ITF451018QNP | ITF451118QNP | ITF451318QNP |
| Ø 5.0 | ITF500718QNP | ITF500818QNP | ITF501018QNP | ITF501118QNP | ITF501318QNP |



**• Regular ( G / H : 2.8 )**

| D \ L | 7 | 8.5 | 10 | 11.5 | 13 |
|---|---|---|---|---|---|
| Ø 3.5 | - | ITF350828QNP | ITF351028QNP | ITF351128QNP | ITF351328QNP |
| Ø 4.0 | ITF400728QNP | ITF400828QNP | ITF401028QNP | ITF401128QNP | ITF401328QNP |
| Ø 4.5 | ITF450728QNP | ITF450828QNP | ITF451028QNP | ITF451128QNP | ITF451328QNP |
| Ø 5.0 | ITF500728QNP | ITF500828QNP | ITF501028QNP | ITF501128QNP | ITF501328QNP |



**• Wide ( G / H : 2.0 )**

| D \ L | 7 | 8.5 | 10 | 11.5 | 13 |
|---|---|---|---|---|---|
| Ø 4.5 | ITFW450720QNP | ITFW450820QNP | ITFW451020QNP | ITFW451120QNP | ITFW451320QNP |
| Ø 5.0 | ITFW500720QNP | ITFW500820QNP | ITFW501020QNP | ITFW501120QNP | ITFW501320QNP |
| Ø 6.0 | ITFW600720QNP | ITFW600820QNP | ITFW601020QNP | ITFW601120QNP | ITFW601320QNP |
| Ø 7.0 | ITFW700720QNP | ITFW700820QNP | ITFW701020QNP | ITFW701120QNP | ITFW701320QNP |



**• Wide ( G / H : 3.0 )**

| D \ L | 7 | 8.5 | 10 | 11.5 | 13 |
|---|---|---|---|---|---|
| Ø 3.5 | ITFW450730QNP | ITFW450830QNP | ITFW451030QNP | ITFW451130QNP | ITFW451330QNP |
| Ø 4.0 | ITFW500730QNP | ITFW500830QNP | ITFW501030QNP | ITFW501130QNP | ITFW501330QNP |
| Ø 4.5 | ITFW600730QNP | ITFW600830QNP | ITFW601030QNP | ITFW601130QNP | ITFW601330QNP |
| Ø 5.0 | ITFW700730QNP | ITFW700830QNP | ITFW701030QNP | ITFW701130QNP | ITFW701330QNP |

# SM I SLA Fixture (No-Mount)



- An implant with an Internal Octagon and 8° Taper Connection structure.
- SLA surface treatment excellent for bone adhesion.
- A body design that is easy to adjust during insertion and has excellent primary stability.
- Corkscrew Thread for a powerful self-threading effect.
- Has Korea's only square-shaped Body Thread, effective for dispersing stress from chewing pressure.
- Surface: SLA.
- Increases convenience as it allows for one-stage surgery instead of two-stage surgery.
- Recommended insertion torque: 40Ncm or less.
- For single molar cases, use a fixture of Ø 4.5mm or larger.

## SLA Fixture

- **Regular ( G / H : 1.8 )**

| D \ L | 7 | 8.5 | 10 | 11.5 | 13 |
|---|---|---|---|---|---|
| Ø 3.5 | - | ITF350818 | ITF351018 | ITF351118 | ITF351318 |
| Ø 4.0 | ITF400718 | ITF400818 | ITF401018 | ITF401118 | ITF401318 |
| Ø 4.5 | ITF450718 | ITF450818 | ITF451018 | ITF451118 | ITF451318 |
| Ø 5.0 | ITF500718 | ITF500818 | ITF501018 | ITF501118 | ITF501318 |

- **Regular ( G / H : 2.8 )**

| D \ L | 7 | 8.5 | 10 | 11.5 | 13 |
|---|---|---|---|---|---|
| Ø 3.5 | - | ITF350828 | ITF351028 | ITF351128 | ITF351328 |
| Ø 4.0 | ITF400728 | ITF400828 | ITF401028 | ITF401128 | ITF401328 |
| Ø 4.5 | ITF450728 | ITF450828 | ITF451028 | ITF451128 | ITF451328 |
| Ø 5.0 | ITF500728 | ITF500828 | ITF501028 | ITF501128 | ITF501328 |

## Cover Screw






- Packing Unit: Cover Screw.
- 1.2 Hex Driver / with 10 Ncm of torque.
- Re-tighten with a torque of 30Ncm after 10 minutes of initial placement.


- **Regular**

| D | CODE |
|---|---|
| Ø 3.5 | PICS3505G |

- **Wide**

| D | CODE |
|---|---|
| Ø 4.3 | PICS6506GA |

## Healing Abutment




- Packing Unit: Healing Abutment.
- Use a 1.2 Hex Driver.
- Recommended tightening torque: 5~8 Ncm (Regular & Wide).
- Re-tighten with a torque of 30Ncm after 10 minutes of initial placement.

**• Regular**

| D \ G/H | 1 | 2 | 3 | 4 | 5 | 7 |
|---|---|---|---|---|---|---|
| Ø 4.8 | PIHA4851G | PIHA4852G | PIHA4853G | PIHA4854G | PIHA4855G | PIHA4857G |

**• Wide**

| D \ G/H | 1 | 2 | 3 | 4 | 5 | 7 |
|---|---|---|---|---|---|---|
| Ø 6.5 | PIHA6571G | PIHA6572G | PIHA6573G | PIHA6574G | PIHA6575G | PIHA6577G |

## SM I Scanbody




| H(mm) | TYPE | CODE | 분류 |
|---|---|---|---|
| 9 | Regular | ISSRP | Non Sub Merged |
| 9 | Wide | ISDWP | Non Sub Merged |

## Fixture Pick-up Impression Coping




- Packing Unit: Impression Coping Body + Guide Pin.
- For impression taking with a Custom Tray.
- 1.2 Hex Driver / with 10 Ncm of torque.

**• Regular**

| D | TYPE \ L | 9 | 12 |
|---|---|---|---|
| Ø 4.8 | OCTA | PICP4807P | PICP4812P |
| | NON-OCTA | PICP4807NP | PICP4812NP |

**• Wide**

| D | TYPE \ L | 9 | 12 |
|---|---|---|---|
| Ø 6.5 | OCTA | PICP6507P | PICP6512P |
| | NON-OCTA | PICP6507NP | PICP6512NP |

## Fixture Transfer Impression Coping




- Packing Unit: Impression Coping Body + Guide Pin.
- For impression taking with a stock tray.
- 1.2 Hex Driver / with 10 Ncm of torque.

**• Regular**

| D | TYPE \ L | 9 | 12 |
|---|---|---|---|
| Ø 4.8 | OCTA | PICT4809P | PICT4812P |
| | NON OCTA | PICT4809NP | PICT4812NP |

**• Wide**

| D | TYPE \ L | 9 | 12 |
|---|---|---|---|
| Ø 6.5 | OCTA | PICT6509P | PICT6512P |
| | NON OCTA | PICT6509NP | PICT6512NP |

## Solid Lab Analog




- Packing Unit: Fixture Analog.
- Represents the intraoral abutment on the working model.

**• Regular**

| D \ L | 4 | 5.5 | 7 |
|---|---|---|---|
| Ø 4.8 | PISA48354 | PISA48355 | PISA48357 |

**• Wide**

| D \ L | 4 | 5.5 | 7 |
|---|---|---|---|
| Ø 6.5 | PISA65434 | PISA65435 | PISA65437 |

## Excellent Solid Lab Analog




- Packing Unit: Fixture Analog.
- Represents the intraoral abutment on the working model.

**• Regular**

| D \ L | 4 | 5.5 | 7 |
|---|---|---|---|
| Ø 4.8 | PIESA48434 | PIESA48435 | PIESA48437 |

## Cemented Abutment



Octa




Non-Octa

- Packing Unit: Abutment + Abutment Screw.
- Use a 1.2 Hex Driver.
- Recommended tightening torque: 30Ncm (Regular & Wide).
- Titanium screw: PIAS2008.
- Scheduled to be changed to All Coating.
- Re-tighten with a torque of 30Ncm after 10 minutes of initial placement.

**• Regular**


| D | L \ | TYPE | CODE |
|---|---|---|---|
| Ø 4.8 | 4 | OCTA | PIECA4304P |
| | | NON-OCTA | PIECA4304NP |
| | 5.5 | OCTA | PIECA4305P |
| | | NON-OCTA | PIECA4305NP |
| | 7 | OCTA | PIECA4307P |
| | | NON-OCTA | PIECA4307NP |

| D | L \ | TYPE | 1 | 2 | 3 | 4 | 5 |
|---|---|---|---|---|---|---|---|
| Ø 4.8 | 5.5 | OCTA | PIECA4815P | PIECA4825P | PIECA4835P | PIECA4845P | PIECA4855P |
| | | NON-OCTA | PIECA4815NP | PIECA4825NP | PIECA4835NP | PIECA4845NP | PIECA4855NP |
| | 6.5 | OCTA | PIECA4816P | PIECA4826P | PIECA4836P | PIECA4846P | PIECA4856P |
| | | NON-OCTA | PIECA4816NP | PIECA4826NP | PIECA4836NP | PIECA4846NP | PIECA4856NP |

**• Wide**

| D | L \ | TYPE | 1 | 2 | 3 | 4 |
|---|---|---|---|---|---|---|
| Ø 6.5 | 5.5 | OCTA | PIECA6515P | PIECA6525P | PIECA6535P | PIECA6545P |
| | | NON-OCTA | PIECA6515NP | PIECA6525NP | PIECA6535NP | PIECA6545NP |

## Excellent Cemented Abutment




- Packing Unit: Abutment + Abutment Screw.
- Use a 1.2 Hex Driver.
- Recommended tightening torque: 30Ncm (Regular & Wide)
- Titanium screw: PIAS2008.
- Scheduled to be changed to All Coating.
- Re-tighten with a torque of 30Ncm after 10 minutes of initial placement.


**· Regular**

| D | TYPE\G/H | 2 | 3 | 4 | 5 |
|---|---|---|---|---|---|
| Ø 6.0 | OCTA | PIECA4625EP | - | - | - |
| | NON-OCTA | PIECA4625ENP | - | - | - |
| Ø 6.5 | OCTA | - | PIECA4635EP | - | - |
| | NON-OCTA | - | PIECA4635ENP | - | - |
| Ø 7.0 | OCTA | - | - | PIECA4845EP | PIECA4855EP |
| | NON-OCTA | - | - | PIECA4845ENP | PIECA4855ENP |

**· Wide**

| D | TYPE\G/H | 2 | 3 | 4 |
|---|---|---|---|---|
| Ø 6.8 | OCTA | PIECA6525EP | - | - |
| | NON-OCTA | PIECA6525ENP | - | - |
| Ø 6.8 | OCTA | - | PIECA6535EP | - |
| | NON-OCTA | - | PIECA6535ENP | - |
| Ø 6.8 | OCTA | - | - | PIECA6545EP |
| | NON-OCTA | - | - | PIECA6545ENP |

## Angled Abutment




- Packing Unit: Abutment + Abutment Screw
- Provides 15° ~ 25° of angulation compensation without the need for trimming.
- Use a 1.2 Hex Driver
- Recommended tightening torque: 20Ncm (Mini), 30Ncm (Regular).
- Titanium screw: PIAS2006.
- Scheduled to be changed to All Coating.
- Re-tighten with a torque of 30Ncm after 10 minutes of initial placement.


**· Regular**

| D | TYPE \ ANGLED | 15° | 25° |
|---|---|---|---|
| Ø 3.5 | OCTA | PIAA3515P | PIAA3525P |

**· Wide**

| D | TYPE \ ANGLED | 15° | 25° |
|---|---|---|---|
| Ø 3.5 | OCTA | PIAA6515P | PIAA6525P |
| | NON-OCTA | PIAA6515NP | PIAA6525NP |

## Milling Abutment




- Packing Unit: Abutment + Abutment Screw.
- Use a 1.2 Hex Driver.
- Recommended tightening torque: 20Ncm (Mini), 30Ncm (Regular).
- Titanium screw: PIAS2008.
- Re-tighten with a torque of 30Ncm after 10 minutes of initial placement.

| D | TYPE \ G/H | 3.0 |
|---|---|---|
| Ø 4.8 | Regular | PIEMA4833P |
| | Regular | PIEMA4833NP |
| Ø 6.5 | Wide | PIEMA6533P |
| | Wide | PIEMA6533NP |

## Excellent Temporary Abutment




- Packing Unit: Abutment + Abutment Screw.
- Used by milling to fabricate temporary prostheses.
- Use a 1.2 Hex Driver.
- Recommended tightening torque: 20Ncm (Regular).
- Titanium screw: PIAS2008.
- Re-tighten with a torque of 30Ncm after 10 minutes of initial placement.

**• Regular**

| D | TYPE \ G/H | 0.0 | 1.0 | 2.0 | 3.0 |
|---|---|---|---|---|---|
| Ø 4.8 | OCTA | PIETA4800P | PIETA4810P | PIETA4820P | PIETA4830P |
| | NON-OCTA | PIETA4800NP | PIETA4810NP | PIETA4820NP | PIETA4830NP |

**• Wide**

| D | TYPE \ G/H | 0.0 | 1.0 | 2.0 | 3.0 |
|---|---|---|---|---|---|
| Ø 6.5 | OCTA | PIETA6500P | PIETA6510P | PIETA6520P | PIETA6530P |
| | NON-OCTA | PIETA6500NP | PIETA6510NP | PIETA6520NP | PIETA6530NP |

## CCM Casting Abutment




- Packing Unit: Casting Abutment + Abutment Screw.
- Used when casting with non-precious metal alloy (Cr-Co-Mo) and fabricating a custom prosthesis.
- Use a 1.2 Hex Driver.
- Recommended tightening torque: 30Ncm (Regular & Wide).
- Titanium screw: PIAS2008.
- Re-tighten with a torque of 30Ncm after 10 minutes of initial placement.

**• Regular**

| D | TYPE | CODE |
|---|---|---|
| Ø 4.8 | OCTA | PICA4800HP |
| | NON-OCTA | PICA4800NP |

**• Wide**

| D | TYPE | CODE |
|---|---|---|
| Ø 6.5 | OCTA | PICA6500HP |
| | NON-OCTA | PICA6500NP |

## Solid Abutment




- Packing Unit: Abutment + Protect Cap.
- An abutment for cement-retained prosthesis fabrication.
- Use a 1.2 Hex Driver.
- Recommended tightening torque: 30Ncm (Regular & Wide)
- Scheduled to be changed to All Coating.
- Re-tighten with a torque of 30Ncm after 10 minutes of initial placement.

**• Regular**

| D \ L | 4 | 5.5 | 7 |
|---|---|---|---|
| Ø 3.5 | PISA**3504**P | PISA**3505**P | PISA**3507**P |

**• Wide**

| D \ L | 4 | 5.5 | 7 |
|---|---|---|---|
| Ø 4.3 | PISA**4304**P | PISA**4305**P | PISA**4307**P |

## Excellent Solid Abutment




- Packing Unit: Abutment + Protect Cap.
- An abutment for cement-retained prosthesis fabrication.
- Use a 1.2 Hex Driver.
- Recommended tightening torque: 30Ncm (Regular)
- Scheduled to be changed to All Coating.
- Re-tighten with a torque of 30Ncm after 10 minutes of initial placement.

**• Regular**

| D | G/H \ L | 4 | 5.5 | 7 |
|---|---|---|---|---|
| Ø 4.3 | 0 | PIESA**4304**P | PIESA**4305**P | PIESA**4307**P |
| Ø 4.8 | 0 | PIESA**4814**P<br>PIESA**4824**P<br>PIESA**4834**P | PIESA**4815**P<br>PIESA**4825**P<br>PIESA**4835**P | PIESA**4817**P<br>PIESA**4827**P<br>PIESA**4837**P |

## Fixture Pin




| CODE | L |
|---|---|
| PIFP3500 | 15 |

## Excellent Solid Abutment




- Packing Unit: Fixture Analog.
- Represents the intraoral fixture on the working model.
- Use separately for Fixture Platform Ø4.8/6.0.

**Size Guide**

- Regular: PIFA4812 (Blue)
- Wide: PIFA6512 (Yellow)

**• Regular**

| D \ L | 12 |
|---|---|
| Ø 4.8 | PIFA**4812** |


**• Wide**

| D \ L | 12 |
|---|---|
| Ø 6.5 | PIFA**6512** |

# Digital Prosthetics

Rite Impression Coping Scanbody
SM II Scanbody
Customized Abutment
Ti-Base
Digital 3D Analog
Cemented Abutment

# Digital Prosthetics



## Bite Impression Coping Scanbody




- Packing Unit : Scan body + Guide Pin.
- Scan Body for Titanium Custom Abutment Fabrication.
- For Model Scan: Long (15mm).
- For Intra-Oral Scan : Short (10mm).
- Fastening by hand using 1.2 Hex Driver.

**Size Classification**

- Guide Pin : Regular(Non), Mini(Brown).
- Body : G/H 2mm(Non), 3mm(Yellow), 4mm(Purple), 6mm(Blue), 8mm(Non) Coloring

• **Mini**

| D I TYPE G/H | 2.0 | 3.0 | 4.0 | 6.0 | 8.0 |
|---|---|---|---|---|---|
| Ø 4.0 Scan body | OBICV**4002**SBP | OBICM**4003**SBP | OBICM**4004**SBP | OBICV**4006**SBP | OBICM**4008**SBP |

• **Regular**

| D I TYPE G/H | 2.0 | 3.0 | 4.0 | 6.0 | 8.0 |
|---|---|---|---|---|---|
| Ø 4.0 Scan body | OBICR**4002**SBP | OBICR**4003**SBP | OBICR**4004**SBP | OBICR**4006**SBP | OBICR**4008**SBP |
| Ø 4.5 Scan body | OBICR**4502**SBP | OBICR**4503**SBP | OBICR**4504**SBP | OBICR**4506**SBP | OBICR**4508**SBP |
| Ø 5.0 Scan body | OBICR**5002**SBP | OBICR**5003**SBP | OBICR**5004**SBP | OBICR**5006**SBP | OBICR**5008**SBP |

## SM II Scanbody




| H(mm) | TYPE | CODE | 2.0 |
|---|---|---|---|
| 12 | Regular | OSBRP | Sub Verged |
| 16 | Mini | OSBMP | Sub Verged |

## Customized Abutment



- Packing Unit: Abutment + Abutment Screw.
- Custom abutments can be fabricated using dental milling equipment.
- Superior fastening precision compared to non-original components.
- Compatible with a wide range of dedicated milling systems.
  (Supported manufacturers: Dyovon, Vatech, Neo, Manix, etc.)
- Re-tighten with a torque of 30Ncm after 10 minutes of initial placement.

• **Mini**

| D | TYPE | CODE | PRODUCT |
|---|---|---|---|
| Ø 10 | Hex | OCMM**1001**HP | ARUMM Imes-Icore |
| | Non-Hex | OCMM**1001**NP | |
| Ø 14 | Hex | OCMM**1401**HP | |
| | Non-Hex | OCMM**1401**NP | |
| Ø 10 | Hex | OCMM**1002**HP | Manix |
| | Non-Hex | OCMM**1002**NP | |
| Ø 14 | Hex | OCMM**1402**HP | |
| | Non-Hex | OCMM**1402**NP | |

• **Regular**

| D | TYPE | CODE | PRODUCT |
|---|---|---|---|
| Ø 10 | Hex | OCMR**1001**HP | ARUMM Imes-Icore |
| | Non-Hex | OCMR**1001**NP | |
| Ø 14 | Hex | OCMR**1401**HP | |
| | Non-Hex | OCMR**1401**NP | |
| Ø 10 | Hex | OCVR**1002**HP | Manix |
| | Non-Hex | OCVR**1002**NP | |
| Ø 14 | Hex | OCVR**1402**HP | |
| | Non-Hex | OCVR**1402**NP | |

## Ti-Base



- Abutment for Cemented, Combination, and Screw-Retained Prosthetics.
- Ti-Base + Zirconia Custom Fabrication Available Using CAD/CAM Equipment.
- Tighten by using 1.2 Hex Driver.
- Recommended Tightening Torque: 20 Ncm (Mini), 30 Ncm (Regular).
- Abutment Screw: OAS1810 (Mini), OAS2006 (Regular).
- Uses Point Implant Library.
- Packing Unit: Abutment + Abutment Screw.
- Re-tighten with a torque of 30Ncm after 10 minutes of initial placement.

### • Mini

| Type | D | H (mm) | G/H / Type | 1.0 | 2.0 | 3.0 | 4.0 |
|---|---|---|---|---|---|---|---|
| Hex | Ø 4.0 | 4.0 | Cylinder | OLPCM4041HP | OLPCM4042HP | OLPCM4043HP | OLPCM4044HP |
| | | 6.0 | | OLPCM4061HP | OLPCM4062HP | OLPCM4063HP | OLPCM4064HP |
| | | 4.0 | Open | OLPOM4041HP | OLPOM4042HP | OLPCM4043HP | OLPOM4044HP |

| Type | D | H (mm) | G/H / Type | 1.0 | 2.0 | 3.0 | 4.0 |
|---|---|---|---|---|---|---|---|
| Non-Hex | Ø 4.0 | 4.0 | Cylinder | OLPCM4041NP | OLPCM4042NP | OLPCM4043NP | OLPCM4044NP |
| | | 6.0 | | OLPCM4061NP | OLPCM4062NP | OLPCM4063NP | OLPCM4064NP |
| | | 4.0 | Open | OLPOM4041NP | OLPOM4042NP | OLPCM4043NP | OLPOM4044NP |

### • Regular

| Type | D | H (mm) | G/H / Type | 1.0 | 2.0 | 3.0 | 4.0 |
|---|---|---|---|---|---|---|---|
| Hex | Ø 4.5 | 4.0 | Cylinder | OLPC4541HP | OLPC4542HP | OLPC4543HP | OLPC4544HP |
| | | 6.0 | | OLPC4561HP | OLPC4562HP | OLPC4563HP | OLPC4564HP |
| | | 4.0 | Open | OLPO4541HP | OLPO4542HP | OLPO4543HP | OLPO4544HP |

| Type | D | H (mm) | G/H / Type | 1.0 | 2.0 | 3.0 | 4.0 |
|---|---|---|---|---|---|---|---|
| Non-Hex | Ø 4.5 | 4.0 | Cylinder | OLPC4541NP | OLPC4542NP | OLPC4543NP | OLPC4544NP |
| | | 6.0 | | OLPC4561NP | OLPC4562NP | OLPC4563NP | OLPC4564NP |
| | | 4.0 | Open | OLPO4541NP | OLPO4542NP | OLPO4543NP | OLPO4544NP |

## Digital 3D Analog



- Intraoral fixture implementation on the working model.
- Tighten by hand using a 1.2 Hex Driver.
- Packing Unit: Digital Lab Analog 1ea.
- Size classification.
  Mini 3.5 (Yellow), Regular 4.0 (Purple)


### • Mini 3.5

| D | CODE |
|---|---|
| Ø 3.5 | DDLA350 |

### • Regular 4.0

| D | CODE |
|---|---|
| Ø 4.0 | DDLA400 |

# Cemented Abutment




- Packing Unit : Abutment + Abutment Screw.
- Loaded with 1.2 Hex Driver.
- Recommended torque: 20Ncm (Mini), 30Ncm (Regular).
- Titanium Screw Catalog Code : OAS1610(Mini), OAS2008(Regular).
- Re-tighten with a torque of 30Ncm after 10 minutes of initial placement.

## • Mini

| D | L | TYPE \ G / H | 1.0 | 2.0 | 3.0 | 4.0 | 5.0 | 6.0 | 7.0 |
|---|---|---|---|---|---|---|---|---|---|
| Ø 4.0 | 4.0 | Hex | OSCAM4041HTP | OSCAM4042HTP | OSCAM4043HTP | OSCAM4044HTP | OSCAM4045HTP | *OSCAM4046HTP | *OSCAM4047HTP |
| | | Non-Hex | OSCAM4041NTP | OSCAM4042NTP | OSCAM4043NTP | OSCAM4044NTP | OSCAM4045NTP | *OSCAM4046NTP | *OSCAM4047NTP |
| | 5.5 | Hex | OSCAM4051HTP | OSCAM4052HTP | OSCAM4053HTP | OSCAM4054HTP | OSCAM4055HTP | *OSCAM4056HTP | *OSCAM4057HTP |
| | | Non-Hex | OSCAM4051NTP | OSCAM4052NTP | OSCAM4053NTP | OSCAM4054NTP | OSCAM4055NTP | *OSCAM4056NTP | *OSCAM4057NTP |
| | 7.0 | Hex | OSCAM4071HTP | OSCAM4072HTP | OSCAM4073HTP | OSCAM4074HTP | OSCAM4075HTP | *OSCAM4076HTP | *OSCAM4077HTP |
| | | Non-Hex | OSCAM4071NTP | OSCAM4072NTP | OSCAM4073NTP | OSCAM4074NTP | OSCAM4075NTP | *OSCAM4076NTP | *OSCAM4077NTP |
| Ø 4.5 | 4.0 | Hex | OSCAM4541HTP | OSCAM4542HTP | OSCAM4543HTP | OSCAM4544HTP | OSCAM4545HTP | OSCAM4546HTP | OSCAM4547HTP |
| | | Non-Hex | OSCAM4541NTP | OSCAM4542NTP | OSCAM4543NTP | OSCAM4544NTP | OSCAM4545NTP | OSCAM4546NTP | OSCAM4547NTP |
| | 5.5 | Hex | OSCAM4551HTP | OSCAM4552HTP | OSCAM4553HTP | OSCAM4554HTP | OSCAM4555HTP | OSCAM4556HTP | OSCAM4557HTP |
| | | Non-Hex | OSCAM4551NTP | OSCAM4552NTP | OSCAM4553NTP | OSCAM4554NTP | OSCAM4555NTP | OSCAM4556NTP | OSCAM4557NTP |
| | 7.0 | Hex | OSCAM4571HTP | OSCAM4572HTP | OSCAM4573HTP | OSCAM4574HTP | OSCAM4575HTP | OSCAM4576HTP | OSCAM4577HTP |
| | | Non-Hex | OSCAM4571NTP | OSCAM4572NTP | OSCAM4573NTP | OSCAM4574NTP | OSCAM4575NTP | OSCAM4576NTP | OSCAM4577NTP |
| Ø 5.0 | 4.0 | Hex | OSCAM5041HTP | OSCAM5042HTP | OSCAM5043HTP | OSCAM5044HTP | OSCAM5045HTP | OSCAM5046HTP | OSCAM5047HTP |
| | | Non-Hex | OSCAM5041NTP | OSCAM5042NTP | OSCAM5043NTP | OSCAM5044NTP | OSCAM5045NTP | OSCAM5046NTP | OSCAM5047NTP |
| | 5.5 | Hex | OSCAM5051HTP | OSCAM5052HTP | OSCAM5053HTP | OSCAM5054HTP | OSCAM5055HTP | OSCAM5056HTP | OSCAM5057HTP |
| | | Non-Hex | OSCAM5051NTP | OSCAM5052NTP | OSCAM5053NTP | OSCAM5054NTP | OSCAM5055NTP | OSCAM5056NTP | OSCAM5057NTP |
| | 7.0 | Hex | OSCAM5071HTP | OSCAM5072HTP | OSCAM5073HTP | OSCAM5074HTP | OSCAM5075HTP | OSCAM5076HTP | OSCAM5077HTP |
| | | Non-Hex | OSCAM5071NTP | OSCAM5072NTP | OSCAM5073NTP | OSCAM5074NTP | OSCAM5075NTP | OSCAM5076NTP | OSCAM5077NTP |


## • Regular

| D | L | TYPE\G / H | 1.0 | 2.0 | 3.0 | 4.0 | 5.0 | 6.0 | 7.0 |
|---|---|---|---|---|---|---|---|---|---|
| Ø 4.0 | 4.0 | Hex | OSCA4041HTP | OSCA4042HTP | OSCA4043HTP | OSCA4044HTP | OSCA4045HTP | OSCA4046HTP | OSCA4047HTP |
| | | Non-Hex | OSCA4041NTP | OSCA4042NTP | OSCA4043NTP | OSCA4044NTP | OSCA4045NTP | OSCA4046NTP | OSCA4047NTP |
| | 5.5 | Hex | OSCA4051HTP | OSCA4052HTP | OSCA4053HTP | OSCA4054HTP | OSCA4055HTP | OSCA4056HTP | OSCA4057HTP |
| | | Non-Hex | OSCA4051NTP | OSCA4052NTP | OSCA4053NTP | OSCA4054NTP | OSCA4055NTP | OSCA4056NTP | OSCA4057NTP |
| | 7.0 | Hex | OSCA4071HTP | OSCA4072HTP | OSCA4073HTP | OSCA4074HTP | OSCA4075HTP | OSCA4076HTP | OSCA4077HTP |
| | | Non-Hex | OSCA4071NTP | OSCA4072NTP | OSCA4073NTP | OSCA4074NTP | OSCA4075NTP | OSCA4076NTP | OSCA4077NTP |
| Ø 4.5 | 4.0 | Hex | OSCA4541HTP | OSCA4542HTP | OSCA4543HTP | OSCA4544HTP | OSCA4545HTP | OSCA4546HTP | OSCA4547HTP |
| | | Non-Hex | OSCA4541NTP | OSCA4542NTP | OSCA4543NTP | OSCA4544NTP | OSCA4545NTP | OSCA4546NTP | OSCA4547NTP |
| | 5.5 | Hex | OSCA4551HTP | OSCA4552HTP | OSCA4553HTP | OSCA4554HTP | OSCA4555HTP | OSCA4556HTP | OSCA4557HTP |
| | | Non-Hex | OSCA4551NTP | OSCA4552NTP | OSCA4553NTP | OSCA4554NTP | OSCA4555NTP | OSCA4556NTP | OSCA4557NTP |
| | 7.0 | Hex | OSCA4571HTP | OSCA4572HTP | OSCA4573HTP | OSCA4574HTP | OSCA4575HTP | OSCA4576HTP | OSCA4577HTP |
| | | Non-Hex | OSCA4571NTP | OSCA4572NTP | OSCA4573NTP | OSCA4574NTP | OSCA4575NTP | OSCA4576NTP | OSCA4577NTP |
| Ø 5.0 | 4.0 | Hex | OSCA5041HTP | OSCA5042HTP | OSCA5043HTP | OSCA5044HTP | OSCA5045HTP | OSCA5046HTP | OSCA5047HTP |
| | | Non-Hex | OSCA5041NTP | OSCA5042NTP | OSCA5043NTP | OSCA5044NTP | OSCA5045NTP | OSCA5046NTP | OSCA5047NTP |
| | 5.5 | Hex | OSCA5051HTP | OSCA5052HTP | OSCA5053HTP | OSCA5054HTP | OSCA5055HTP | OSCA5056HTP | OSCA5057HTP |
| | | Non-Hex | OSCA5051NTP | OSCA5052NTP | OSCA5053NTP | OSCA5054NTP | OSCA5055NTP | OSCA5056NTP | OSCA5057NTP |
| | 7.0 | Hex | OSCA5071HTP | OSCA5072HTP | OSCA5073HTP | OSCA5074HTP | OSCA5075HTP | OSCA5076HTP | OSCA5077HTP |
| | | Non-Hex | OSCA5071NTP | OSCA5072NTP | OSCA5073NTP | OSCA5074NTP | OSCA5075NTP | OSCA5076NTP | OSCA5077NTP |
| Ø 5.5 | 4.0 | Hex | OSCA5541HTP | OSCA5542HTP | OSCA5543HTP | OSCA5544HTP | OSCA5545HTP | OSCA5546HTP | OSCA5547HTP |
| | | Non-Hex | OSCA5541NTP | OSCA5542NTP | OSCA5543NTP | OSCA5544NTP | OSCA5545NTP | OSCA5546NTP | OSCA5547NTP |
| | 5.5 | Hex | OSCA5551HTP | OSCA5552HTP | OSCA5553HTP | OSCA5554HTP | OSCA5555HTP | OSCA5556HTP | OSCA5557HTP |
| | | Non-Hex | OSCA5551NTP | OSCA5552NTP | OSCA5553NTP | OSCA5554NTP | OSCA5555NTP | OSCA5556NTP | OSCA5557NTP |
| | 7.0 | Hex | OSCA5571HTP | OSCA5572HTP | OSCA5573HTP | OSCA5574HTP | OSCA5575HTP | OSCA5576HTP | OSCA5577HTP |
| | | Non-Hex | OSCA5571NTP | OSCA5572NTP | OSCA5573NTP | OSCA5574NTP | OSCA5575NTP | OSCA5576NTP | OSCA5577NTP |
| Ø 6.0 | 4.0 | Hex | OSCA6041HTP | OSCA6042HTP | OSCA6043HTP | OSCA6044HTP | OSCA6045HTP | OSCA6046HTP | OSCA6047HTP |
| | | Non-Hex | OSCA6041NTP | OSCA6042NTP | OSCA6043NTP | OSCA6044NTP | OSCA6045NTP | OSCA6046NTP | OSCA6047NTP |
| | 5.5 | Hex | OSCA6051HTP | OSCA6052HTP | OSCA6053HTP | OSCA6054HTP | OSCA6055HTP | OSCA6056HTP | OSCA6057HTP |
| | | Non-Hex | OSCA6051NTP | OSCA6052NTP | OSCA6053NTP | OSCA6054NTP | OSCA6055NTP | OSCA6056NTP | OSCA6057NTP |
| | 7.0 | Hex | OSCA6071HTP | OSCA6072HTP | OSCA6073HTP | OSCA6074HTP | OSCA6075HTP | OSCA6076HTP | OSCA6077HTP |
| | | Non-Hex | OSCA6071NTP | OSCA6072NTP | OSCA6073NTP | OSCA6074NTP | OSCA6075NTP | OSCA6076NTP | OSCA6077NTP |
| Ø 6.5 | 4.0 | Hex | OSCA6541HTP | OSCA6542HTP | OSCA6543HTP | OSCA6544HTP | OSCA6545HTP | OSCA6546HTP | OSCA6547HTP |
| | | Non-Hex | OSCA6541NTP | OSCA6542NTP | OSCA6543NTP | OSCA6544NTP | OSCA6545NTP | OSCA6546NTP | OSCA6547NTP |
| | 5.5 | Hex | OSCA6551HTP | OSCA6552HTP | OSCA6553HTP | OSCA6554HTP | OSCA6555HTP | OSCA6556HTP | OSCA6557HTP |
| | | Non-Hex | OSCA6551NTP | OSCA6552NTP | OSCA6553NTP | OSCA6554NTP | OSCA6555NTP | OSCA6556NTP | OSCA6557NTP |
| | 7.0 | Hex | OSCA6571HTP | OSCA6572HTP | OSCA6573HTP | OSCA6574HTP | OSCA6575HTP | OSCA6576HTP | OSCA6577HTP |
| | | Non-Hex | OSCA6571NTP | OSCA6572NTP | OSCA6573NTP | OSCA6574NTP | OSCA6575NTP | OSCA6576NTP | OSCA6577NTP |
| Ø 7.0 | 4.0 | Hex | OSCA7041HTP | OSCA7042HTP | OSCA7043HTP | OSCA7044HTP | OSCA7045HTP | OSCA7046HTP | OSCA7047HTP |
| | | Non-Hex | OSCA7041NTP | OSCA7042NTP | OSCA7043NTP | OSCA7044NTP | OSCA7045NTP | OSCA7046NTP | OSCA7047NTP |
| | 5.5 | Hex | OSCA7051HTP | OSCA7052HTP | OSCA7053HTP | OSCA7054HTP | OSCA7055HTP | OSCA7056HTP | OSCA7057HTP |
| | | Non-Hex | OSCA7051NTP | OSCA7052NTP | OSCA7053NTP | OSCA7054NTP | OSCA7055NTP | OSCA7056NTP | OSCA7057NTP |
| | 7.0 | Hex | OSCA7071HTP | OSCA7072HTP | OSCA7073HTP | OSCA7074HTP | OSCA7075HTP | OSCA7076HTP | OSCA7077HTP |
| | | Non-Hex | OSCA7071NTP | OSCA7072NTP | OSCA7073NTP | OSCA7074NTP | OSCA7075NTP | OSCA7076NTP | OSCA7077NTP |

# MUA

Multi Straight Abutment
Multi Angled Abutment
Multi Abutment Driver
Multi Healing Cap
Multi Scan body
Multi Tranfer Impression Coping
Multi Lab Analog
Multi Impression Coping
Ti-Base
Multi Combination Cylinder
Multi Temporary Cylinder
Multi Casting Cylinder
Multi Plastic Cylinder
Ti-Base Screw
Multi Cylinder Screw

# MUA Flowchart



## Multi Straight Abutment






Multi Abutment Carrier

* Multi Abutment (Multi Abutment Carrier(OMCR480C)


• Mini

| D\ G/H | 1.0 | 2.0 | 3.0 | 4.0 | 5.0 |
|---|---|---|---|---|---|
| Ø 4.8 | OMAM4801TP | OMAM4802TP | OMAM4803TP | OMAM4804TP | OMAM4805TP |

• Regular

| D\ G/H | 1.0 | 2.0 | 3.0 | 4.0 | 5.0 |
|---|---|---|---|---|---|
| Ø 4.8 | OMA4801TP | OMA4802TP | OMA4803TP | OMA4804TP | OMA4805TP |

## Multi Angled Abutment






Multi Abutment Carrier



Multi Angled Abutment Screw

* Multi Angled Abutment=Multi Abutment Carrier(OMCR1400) + Multi Angled Abutment Screw(OMAS2007)

• Mini

| D | Angle\G/H | 1.0 | 2.0 | 3.0 | 4.0 | 5.0 |
|---|---|---|---|---|---|---|
| Ø 4.8 | 17° | OMAAM480217AP | OMAAM480317AP | - | OMAAM480417AP | - |
| | 30° | - | - | OMAAM480330AP | OMAAM480430AP | OMAAM480530AP |

• Regular

| D | Angle\G/H | 2.5 | 3.0 | 3.5 | 4.0 | 5.0 |
|---|---|---|---|---|---|---|
| Ø 4.8 | 17° | OMAA480217AP | OMAA480317AP | - | OMAA480417AP | OMAA480517AP |
| | 30° | - | - | OMAA480330AP | OMAA480430AP | OMAA480530AP |

## Multi Abutment Driver



| CODE |
| --- |
| OVDM3300 |

## Multi Healing Cap



| CODE |
| --- |
| OMHC4806N |

## Multi Scan body



| CODE | HEX |
| --- | --- |
| OMSB4806N | Non-Hex |

## Multi Tranfer Impression Coping




| CODE |
| --- |
| OMICT4806N |

## Multi Lab Analog



| CODE |
| --- |
| OMLA480 |

## Multi Impression Coping





* Multi Impression Coping+Multi Guide pin(OVGP1415)

| L / Type | Hex | Non-Hex |
|---|---|---|
| 8 | OMICP4808HP | OMICP4808NP |

• **Multi Guide pin**

| L / Type | 3 | 8 | 10 | 13 |
|---|---|---|---|---|
| Hex | OMGP1410 | OMGP1415 | OMGP1417 | OVGP1420 |

## Ti-Base MUA




* Multi Ti Base+Multi Ti Base Screw(OVTBS1404)

| L | Angle / H | Hex | Non-Hex |
|---|---|---|---|
| 4.3 | 5° | OVTB0405NP | - |
| 4.3 | 5° | - | OMTB0505NP |
| 4.3 | 10° | OMTB04*0NP | - |

## Multi Combination Cylinder



* Multi Temporary cylinder+Multi Cylinder Screw(OVAS1404)

| L / Type | Hex | Non-Hex |
|---|---|---|
| 7.3 | OMCC4807HP | OMCC4807NP |

## Multi Temporary Cylinder



* Multi Abutment+Multi Abutment Carrier(OMCR4800)

| L / Type | Hex | Non-Hex |
|---|---|---|
| 12 | OMTC48*2HP | OMTC4812NP |

## Multi Casting Cylinder



* Multi Casting cylinder+Multi Cylinder Screw (OMAS1404)

| L / Type | Hex | Non-Hex |
|---|---|---|
| 12 | OMCAC4812HP | OMCAC4812NP |


## Multi Plastic Cylinder



* Multi Plastic cylinder+Multi Cylinder Screw (OMAS1404)

| L / Type | Hex | Non-Hex |
|---|---|---|
| 12 | OMPC4812HP | OMPC4812NP |

## Ti-Base Screw



| CODE |
|---|
| OMTBS1404 |

## Multi Cylinder Screw




| CODE |
|---|
| OMAS1404 |

# KIT

Point Surgical Full Kit (OTF-KIT)
Point Surgical Standard Kit (OTA-KIT)
Point Surgical Compact Kit (OTS-KIT)
Point NaVi Kit
Point Narrow NaVi Kit
Drilling protocol
Point Prosthetic Kit
Bite Impression Kit
Point Full S.S.T KIT
Point Simple S.S.T KIT
Point S.O.S KIT
Bone Expander KIT
Sinus Crestal Approach KIT
Sinus Lateral Approach KIT
Implant & Screw Remover KIT(s)
Simple Screw Remover KIT

# Point Surgical Full Kit

(OTF-KIT) **For SM II, TN**




**LANCE DRILL**

CGD1300

**Ø2.3 TWIST DRILL**

CID2308
CID2310
CID2313

**Ø3.0 TWIST DRILL**

CID27 5

**TAPER DRILL F 3.5**

OTFDA3507
OTFDA3508
OTFDA3510
OTFDA3511
OTFDA3513

**TAPER DRILL F 4.0**

OTFDA4007
OTFDA4008
OTFDA4010
OTFDA4011
OTFDA4013

**TAPER DRILL F 4.5**

OTFDA4507
OTFDA4508
OTFDA4510
OTFDA4511
OTFDA4513

**TAPER DRILL F 5.0**

OTFDA5007
OTFDA5008
OTFDA5010
OTFDA5011
OTFDA5013

**HAND DRIVER**

CHDS1200
CHDL1200

**TORQUE DRIVER**

OTDS1200
OTDL1200

**NO-MOUNT DRIVER**

OFDM25005
OFDM2500
OFDM2700

**TAPER DRILL F 5.5**

OTFDA5507
OTFDA5508
OTFDA5510
OTFDA5511
OTFDA5513

**TAPER DRILL F 6.0**

OTFDA6007
OTFDA6008
OTFDA6010
OTFDA6011
OTFDA6013

**TAPER DRILL F 6.5**

OTFDA6507
OTFDA6508
OTFDA6510
OTFDA6511
OTFDA6513

**TAPER DRILL F 7.0**

OTFDA7007
OTFDA7008
OTFDA7010
OTFDA7011
OTFDA7013

**SIDECUT DRILL**

OSCD2000
OSCD2000

**DRILL EXTENSION**

PDE2300

**PARALLEL PIN**

OPP4000
OPP5000

**FLXTURE DRIVER**

OFDT2000S
OFDT2000
OFDT2100

**MOUNT DRIVER**

OSMDS
OSMDL

**TORQUE EXTENSION**

OSTE

**TAPER CORTICAL DRILL**

OTCD3000
OTCD3500
OTCD4000
OTCD4500
OTCD5000
OTCD5500
OTCD6000
OTCD6500
OTCD7000

**DEPTH GAUGE**


OSDG

**TORQUE WRENCH**

OTW

# Point Surgical Standard Kit

(OTA-KIT) **For SM II, TN**




**LANCE DRILL**

OSD1900

**Ø2.3 TWIST DRILL**

CID2308
CID2310
CID2312

**Ø3.0 TWIST DRILL**

CID2710

**TAPER DRILL F 3.5**

OTFDA3507
OTFDA3508
OTFDA3510
OTFDA3511
OTFDA3512

**TAPER DRILL F 4.0**

OTFDA4007
OTFDA4008
OTFDA4010
OTFDA4011
OTFDA4013

**TAPER DRILL F 4.5**

OTFDA4507
OTFDA4508
OTFDA4510
OTFDA4511
OTFDA4513

**TAPER DRILL F 5.0**

OTFDA5007
OTFDA5008
OTFDA5010
OTFDA5011
OTFDA5013

**SIDECUT DRILL**

OSCD2000
OSCD3000

**DRILL EXTENSION**

FDE2300

**PARALLEL PIN**

OPP4000
OPP5000

**FLXTURE DRIVER**

OFDT2500S
OFDT2500
OFDT2100

**MOUNT DRIVER**

OSMDS
OSMDL

**MOUNT EXTENSION**

OMETS
OMETL

**HAND DRIVER**

OHDS1200
OHDL1200

**TORQUE DRIVER**

OTDS1200
OTDL1200

**NO-MOUNT DRIVER**

OFDM2500S
OFDM2500
OFDM2100

**MOUNT REMOVAL TOOL**

OR1M
OR1R

**TORQUE EXTENSION**

OSTE

**TAPER CORTICAL DRILL**

OTCD3000
OTCD3500
OTCD4000
OTCD4500
OTCD5000

**DEPTH GAUGE**


OGDG

**TORQUE WRENCH**

OTW

# Point Surgical Compact Kit

(OTS-KIT)




**LANCE DRILL**

OGD1800

**Ø2.3 TWIST DRILL**

OID2370

**Ø3.0 TWIST DRILL**

OID2715

**TAPER DRILL F 3.5**

OTFD3506
OTFD3510
OTFD3511
OTFD3513

**TAPER DRILL F 4.0**

OTFD4009
OTFD4010
OTFD4011
OTFD4013

**TAPER DRILL F 4.5**

OTFD4506
OTFD4510
OTFD4511
OTFD4513

**TAPER DRILL F 5.0**

OTFD5009
OTFD5010
OTFD5011
OTFD5013

**SIDECUT DRILL**

OSCD2000

**DRILL EXTENSION**

OCE2300

**PARALLEL PIN**

OPP4000
OPP5000

**TORQUE EXTENSION**

OSTE

**NO-MOUNT DRIVER**

OFDM05035
OFDM05100
OFDM02100

**FIXTURE DRIVER**

OFDT25CCS
OFDT25CC
OFDT21CC

**HAND DRIVER**

O-DS1200
O-DL1200

**TORQUE DRIVER**

OTDS1200
OTDL1200

**TORQUE WRENCH**

OTW

## Lance Drill



- Drill used to mark the implantation location.
- Sharp cutting edge for easy positioning.

| CODE |
|---|
| OGD1900 |

## Side Cut Drill (Lindemann)



- Drill used when expanding drill size.
- Drill cutting edge allows side cutting.

| D | CODE |
|---|---|
| Ø 2.0 | OSCD2000 |
| Ø 3.0 | OSCD3000 |

## Twist Drill



- Drill used with a guide.
- Ø 2.3 Initial drill.
- Ø 3.0 Drill with a pilot function to minimize splashing.

* Ø 2.7 Drill is not included in KIT. (available for individual orders only)

| D/L | 7 | 8.5 | 10 | 11.5 | 13 | 15 | 18 |
|---|---|---|---|---|---|---|---|
| Ø 2.3 | OID2307 | OID2308 | OID2310 | OID2311 | OID2313 | - | - |
| Ø 2.7 | - | - | - | - | - | OID2715 | OID2718 |
| Ø 3.0 | - | - | - | OID3011 | - | OID3015 | - |

## Taper Drill



- Drills dedicated for taper type fixtures by diameter and length accordingly.
- Stopper Drill with 1mm extra length.
- Color of handle indicates fixture diameter.

• **Stopper**

| D/L | 7.0 | 8.5 | 10 | 11.5 | 13 |
|---|---|---|---|---|---|
| F 3.5 | OTFD3507 | OTFD3508 | OTFD3510 | OTFD3511 | OTFD3513 |
| F 4.0 | OTFD4007 | OTFD4008 | OTFD4010 | OTFD4011 | OTFD4013 |
| F 4.5 | OTFD4507 | OTFD4508 | OTFD4510 | OTFD4511 | OTFD4513 |
| F 5.0 | OTFD5007 | OTFD5008 | OTFD5010 | OTFD5011 | OTFD5013 |
| F 6.0 | OTFD6007 | OTFD6008 | OTFD6010 | OTFD6011 | OTFD6013 |
| F 7.0 | OTFD7007 | OTFD7008 | OTFD7010 | OTFD7011 | OTFD7013 |

## Taper Drill



- Drills dedicated for taper type fixtures by diameter and length accordingly.
- Stopper Drill with 1mm extra length.
- Color of handle indicates fixture diameter.

• **Stopper**

| D/L | 6.0 | 7.0 | 8.5 | 10 | 11.5 | 13 |
|---|---|---|---|---|---|---|
| F 3.5 | - | OTFDA3507 | OTFDA3508 | OTFDA3510 | OTFDA3511 | OTFDA3513 |
| F 4.0 | OTFDA4006 | OTFDA4007 | OTFDA4008 | OTFDA4010 | OTFDA4011 | OTFDA4013 |
| F 4.5 | OTFDA4506 | OTFDA4507 | OTFDA4508 | OTFDA4510 | OTFDA4511 | OTFDA4513 |
| F 5.0 | OTFDA5006 | OTFDA5007 | OTFDA5008 | OTFDA5010 | OTFDA5011 | OTFDA5013 |
| F 5.5 | OTFDA5506 | OTFDA5507 | OTFDA5508 | OTFDA5510 | OTFDA5511 | OTFDA5513 |
| F 6.0 | OTFDA6006 | OTFDA6007 | OTFDA6008 | OTFDA6010 | OTFDA6011 | OTFDA6013 |
| F 6.5 | OTFDA6506 | OTFDA6507 | OTFDA6508 | OTFDA6510 | OTFDA6511 | OTFDA6513 |
| F 7.0 | OTFDA7006 | OTFDA7007 | OTFDA7008 | OTFDA7010 | OTFDA7011 | OTFDA7013 |

## Taper Cortical Drill



- Used when removing cortical bone to decrease the implantation torque on hard bones.
- Equipped with a dedicated Drill for each fixture diameter.
- It is recommended to drill to the bottom of the marking line.

* F5.5, F6.0, F6.5, F7.0 not included in KIT (available only for individual order)

| SPEC | CODE |
|---|---|
| F 3.0 | OTCD3000 |
| F 3.5 | OTCD3500 |
| F 4.0 | OTCD4000 |
| F 4.5 | OTCD4500 |
| F 5.0 | OTCD5000 |
| F 6.0 | OTCD6000 |
| F 7.0 | OTCD7000 |

## No-Mount Driver




- Driver to place a fixture without a mount.
- Recommended rpm: 50 or less.

* When removing the No-Mount Driver, be sure to disengage vertically.
(If not separated vertically, the risk of fracture is high)

* Extra Long : 5mm longer than Long Size.

* Products ordered separately: OFDM2100, OFDM2100S, OFDM2100EL, OFDM2500EL.

- **Mini**

| TYPE | CODE |
|---|---|
| S | OFDM2100S |
| L | OFDM2100 |
| EL | OFDM2100EL |

- **Regular**

| TYPE | CODE |
|---|---|
| S | OFDM2500S |
| L | OFDM2500 |
| EL | OFDM2500EL |

- **SM I Type**

| TYPE | CODE |
|---|---|
| S | OFDM3100S |
| L | OFDM3100L |

## Fixture Driver



- Driver for placing the fixture by hand.
- Has a guide, so the fixture fastening improved.

* When removing the No-Mount Driver, be sure to disengage vertically.
(If not separated vertically, the risk of fracture is high)

- **Mini**

| TYPE | CODE |
|---|---|
| T/L | OFDT2100 |

- **Regular**

| TYPE | CODE |
|---|---|
| T/S | OFDT2500S |
| T/L | OFDT2500 |

## Hand Driver



- 1.2 Hex Driver.
- A driver for tightening or removing abutments, screws, etc.

| TYPE | CODE |
|---|---|
| Hand Torque1.2ES | OHDS1207 |
| Hand Torque1.2S | OHDS1200 |
| Hand Torque1.2L | OHDL1200 |

※ OHDS1207 : 7mm long Hex Driver

## Torque Driver



- 1.2 Hex Torque Driver
- Torque driver for abutment and screw.
- Maximum Torque: 60 Ncm

| TYPE | CODE |
|---|---|
| Torque1.2ES | OTDS1207 |
| Torque1.2S | OTDS1200 |
| Torque1.2L | OTDL1200 |

※ OTDS1207 : 7mm Long Torque Hex Driver

## Drill Extension



- An instrument that extends the Length of drills and other instrument used with implant motor.
- After inserting the drill into the drill extension, lightly press the drill until it clicks. It tightens by turning.

| CODE |
|---|
| PDE2300 |

## Parallel Pin



- A tool used to check whether the drilling hole is drilled at the desired angle after initial drilling.
- Pein check available after Ø 2.2 and Ø 3.0 drilling.

| TYPE | CODE |
|---|---|
| Ø 4.0 | OPP4000 |
| Ø 5.0 | OPP5000 |

## Torque Extension



- Tool that extends (+10mm) the Length of Torque Ratchet tool.
- Mainly usec when interference with adjacent teeth occurs.
- Maximum recommended torque: 40 Ncm

| CODE |
|---|
| OSTE |

## Torque Wrench



- A wrench used to adjust the implant placement position or to tighten the abutment, screw, etc. with a constant torque.
- Torque application of 10, 20, 30Ncm is possible, do not use more than 40Ncm.
- Pull the bar and apply torque by aligning the center of the bar with desired torque value.

| CODE |
|---|
| OTW |

## Depth Gauge



- Instrument for measuring drilling depth


| CODE |
|---|
| OSDG |

# Point NAVI Kit

(Guided Implant Kit)

**CODE**

NAVI-KIT




**Ø3.0 STARAIGHT DRILL**

TSID3007_LEDT
TSID3008_LEDT
TSID3010_LEDT
TSID3011_LEDT
TSID3013_LEDT

**Ø3.5 TAPER DRILL**

TSFD3507_LEDT
TSFD3508_LEDT
TSFD3510_LEDT
TSFD3511_LEDT
TSFD3513_LEDT

**Ø4.0 TAPER DRILL**

TSFD4007_LEDT
TSFD4008_LEDT
TSFD4010_LEDT
TSFD4011_LEDT
TSFD4013_LEDT

**Ø4.5 TAPER DRILL**

TSFD4507_LEDT
TSFD4508_LEDT
TSFD4510_LEDT
TSFD4511_LEDT
TSFD4513_LEDT

**Ø5.0 TAPER DRILL**

TSFD5007_LEDT
TSFD5008_LEDT
TSFD5010_LEDT
TSFD5011_LEDT
TSFD5013_LEDT

**NO-MOUNT DRIVER**

AOFDHM_LEDT
AOFDHR_LEDT

**FIXTURE DRIVER HAND**

AOFDTM_LEDT
AOFDTR_LEDT

**TISSUE PUNCH**

AGTP_LEDT

**POINT DRILL**

AGPD_LEDT

**BONT PROFILE**

AGBP_LEDT

**HAND DRIVER**

PHD1203

**TUBE DRILL**

TSID2210_LEDT


**DRILL GUIDE TUBE**

AGTD_LEDT

**TORQUE WRENCH**

TW

## Tissue Punch




- Instrument used for flapless procedure.
- Combined use for Soft Tissue Punching and Hard Tissue Fitting.
- Used for bone leveling.
- It is marked at 1.5mm intervals, so the drilling depth can be adjusted according to sleeve offset.

| CODE |
| --- |
| AGTP_LEDT |

## Point Drill



- Recommended RPM: 800.
- A drill that forms a hole in the bone for easy initial drilling.
- 9mm bone cutting based on sleeve top.
- Easy access to posterior teeth for patients with small mouth openings.
- Do not use on sharp ridge (use 2.2 Drill Tube).

| CODE |
| --- |
| AGPD_LEDT |

## Tube Drill




- Used when the drill is expected to slip on a sharp ridge.
- Used when precise drilling is required in thin bones with narrow bone width.
- Be careful when drilling less than 10mm is required.
- Use with Tube.

| CODE |
| --- |
| TSID2210_LEDT |


## Drill Guide Tube




- Guide role of Tube Drill.

| CODE |
| --- |
| AGTD_LEDT |

# Guide Taper Drill






- 7.0, 8.5, 10, 11.5, Stopper Drill
- 13mm Non Stopper Drill
- Drilling is possible up to a drill that fits the fixture size.
- The color coding of the shaft indicates the diameter and corresponding fixture diameter.

| D / L | 7.0 | 8.5 | 10 | 11.5 | 13 |
|---|---|---|---|---|---|
| ⌀ 3.0 | TSID3007_LEDT | TSID3008_LEDT | TSID3010_LEDT | TSID3011_LEDT | TSID3013_LEDT |
| ⌀ 3.5 | TSID3507_LEDT | TSID3508_LEDT | TSID3510_LEDT | TSID3511_LEDT | TSID3513_LEDT |
| ⌀ 4.0 | TSID4007_LEDT | TSID4008_LEDT | TSID4010_LEDT | TSID4011_LEDT | TSID4013_LEDT |
| ⌀ 4.5 | TSID4507_LEDT | TSID4508_LEDT | TSID4510_LEDT | TSID4511_LEDT | TSID4513_LEDT |
| ⌀ 5.0 | TSID5007_LEDT | TSID5008_LEDT | TSID5010_LEDT | TSID5011_LEDT | TSID5013 _LEDT |

## Fixture Driver - HP Hand Driver



- A tool for placing fixtures using the engine low speed handpiece.
- Designed for Mini Fixture (Yellow code) and Regular Fixture (Green code).
- It is marked at 1.5mm intervals, so the depth of installation can be adjusted according to the sleeve offset.

| TYPE | CODE |
|---|---|
| MINI | AGFDHV_LEDT |
| REGULAR | AGFDHR_LEDT |

## Fixture Driver - HD Hand Driver



- A tool for final placement of fixture using a wrench (Fixture depth, Hex direction control).
- Designed for Mini Fixture (Yellow code) and Regular Fixture (Green code).
- It is marked at 1.5mm intervals, so the depth of installation can be adjusted according to the sleeve offset.

| TYPE | CODE |
|---|---|
| MINI | AGFDTML_LEDT |
| REGULAR | AGFDTR_LEDT |




## Bone Profile



- Used to remove the bone created around the fixture during the 1st and 2nd surgery stages.
- Depending on the connection of the fixture, remove the bone around connection and use it before connecting the healing abutment or other abutment.
- There is a guide part, so it is easy to connect the fixture, and the hex and screw inside the fixture are protected.

| CODE |
|---|
| AGBP_LEDT |


# Point Narrow NAVI Kit

(Guided Implant Kit)

**CODE**

NARROW NAVI-KIT

- Guide Kit used in narrow bone ridge.
- Connection : Mini.
- Compatible with Nobel and Straumann




**Ø2.2 DRILL**



GTD2208 LED1
GTD2210 LED1
GTD2212 LED1
GTD2216 LED1

**Ø2.8 DRILL**



GTD2808 LED1
GTD2810 LED1
GTD2812 LED1
GTD2816 LED1

**Ø3.4 DRILL**



GTD3408 LED1
GTD3410 LED1
GTD3412 LED1
GTD3416 LED1

**TISSUE PUNCH**



ACTPJ EDT

**POINT DRILL**



GPD2404 LED1

**OSSTEM T/W**



FDOM_J EDT

**OSSTEM H/P**



FDOMHP_J EDT

**NOBEL T/W**



FDNM_J EDT

**STRAUMANN T/W**




FDSM_J EDT

**DRILL GUIDE TUBE**



ACTG_J EDT

## Tissue Punch



- Instrument used for flapless procedure.
- Combined use for Soft Tissue Punching and Hard Tissue Flatting.
- Used for bone leveling.
- It is marked at 1.5mm intervals, so the drilling depth can be adjusted according to sleeve offset.

| CODE |
| --- |
| AGTP_LEDT |

## Point Drill



- Recommended RPM: 800.
- A drill that forms a hole in the bone for easy initial drilling.
- 8mm bone cutting based on sleeve top.
- Easy access to posterior teeth for patients with small mouth openings.
- Do not use on sharp ridge (use 2.2 Drill Tube).

| CODE |
| --- |
| GPD2704_LEDT |


## Drill Guide Tube




- Guide role of Tube Drill.

| CODE |
| --- |
| AGTD_LEDT |

## Guide Taper Drill




- 8.5, 10, 11.5, Stopper Drill.
- 13~16mm Non Stopper Drill.
- Drilling is possible up to a drill that fits the fixture size 1.

| D / L | 8.5 | 8.5 | 8.5 | 8.5 |
|---|---|---|---|---|
| Ø 2.2 | GTD22C8_LEDT | GTD2210_LEDT | GTD2212_LEDT | GTD2216_LEDT |
| Ø 2.8 | GTD28C8_LEDT | GTD2810_LEDT | GTD2812_LEDT | GTD2816_LEDT |
| Ø 3.4 | GTD34C8_LEDT | GTD3410_LEDT | GTD3412_LEDT | GTD3416_LEDT |

## Fixture Driver Contra Type - HP Hand Driver



- A tool for placing fixtures using the engine low speed handpiece.
- Designed for Mini Fixture (Yellow marking) and Regular Fixture (Green marking).
- It is marked at 1.5mm intervals, so the depth of installation can be adjusted according to the sleeve offset.

| TYPE | CODE |
|---|---|
| MINI | FDCMHP_LEDT |

## Fixture Driver – HD Hand Driver




- A tool for final placement of fixture using a wrench (Fixture depth, Hex direction control).
- Designed for SMII Mini Fixture (Compatible with Osstem).
- Compatible with Straumann, Nobelbiocera.
- It is marked at 1.5mm intervals, so the depth of installation can be adjusted according to the sleeve offset.


| TYPE | CODE |
|---|---|
| SMI, OSSTEM TS III MINI | AGFDTM_LEDT |
| STRUMANN | AGFDTR_LEDT |
| NOBELBIOCERA | |

# Drilling protocol


## Ø4.0 X 10 + 0.0mm (Offset)




## Ø4.0 X 10 + 1.5mm (Offset)




## Ø4.0 X 10 + 3.0mm (Offset)




## Ø4.0 X 10 + 3.0mm (Offset)




## Ø3.0 X 10 + 0.0mm
(Offset)




## Ø3.0 X 10 + 1.5mm
(Offset)





## Ø3.0 X 10mm + 3.0mm
(Offset)



# Point Prosthetics KIT

**CODE**

PRO KIT




**TORQUE DRIVER**
CTDS1200
CTDL1200

**HAND DRIVER**
CHDS1207
CHDS1200
CHDL1200

**MACHINE DRIVER**
OVDS1200
OVDL1200


**SM1 SOLID DRIVER**
OSDS

**SM2 SOLID DRIVER**
ORDS4000
ORDL4000

**SLOT DRIVER**
OTDS0500
OVDS0500

**TORQUE WRENCH**
DTW

## Torque Driver




- 1.2 Hex Torque Driver.
- Torque driver for abutment and screw.
- Maximum Torque: 60 Ncm.

| D | L | CODE |
|---|---|---|
| Ø2 | 25 | OTDL1225 |
| Ø2 | 20 | OTDL1200 |
| Ø2 | 15 | OTDS1215 |
| Ø2 | 13 | OTDS1200 |
| Ø2 | 7 | OTDS1207 |
| Ø1.6 | 20 | OTSDL1200 |
| Ø1.6 | 13 | OTSDS1200 |

## Hand Driver



- 1.2 Hex Driver.
- A driver for tightening or removing abutments, screws, etc.

| TYPE | CODE |
|---|---|
| Hand Torque1.2SS | CHDS1207 |
| Hand Torque1.2S | CHDS1200 |
| Hand Torque1.2L | CHDL1200 |

* CHDS1207 : Length 7mm Hex Driver

## Machine Driver




- 1.2 Hex Machine Driver.
- Machine driver applying torque to abutment and screw.

| TYPE | CODE | L |
|---|---|---|
| Machine Torque1.2ES | OMDSS1200 | 17 |
| Machine Torque1.2S | OMDS1200 | 23 |
| Machine Torque1.2L | OMDL1200 | 29 |

## SM1 Solid Driver



- Driver dedicated for SM1 Solid Abutment.
- Insert the abutment's groove into the driver's triangular marking part and apply torque.

| TYPE | CODE |
|---|---|
| Solid driver Short | OSDS |
| Solid driver Long | OSDL |

## SM2 Solid Driver



- Driver dedicated for SM2 Solid Abutment.
- Insert the abutment's groove into the driver's triangular marking part and apply torque.

| TYPE | CODE |
|---|---|
| Solid driver Short Φ4.0 | ORDS4000 |
| Solid driver Long Φ4.0 | ORDL4000 |

## Torque Wrench



- A wrench used to adjust the implant placement position or to tighten the abutment, screw, etc. with a constant torque.
- Torque application of 10, 20, 30Ncm is possible, do not use more than 40Ncm.
- Pull the bar and apply torque by aligning the center of the bar with desired torque value.

| CODE |
|---|
| OTW |

## Slot Driver

A tool used on the slot when the hex of a Healing Abutment is damaged and force cannot be applied with a 1.2 Hex Driver.




| L | CODE |
|---|---|
| 25 | OTDEL0500 |
| 20 | OTDL0500 |
| 13 | OTDS0500 |




| L | CODE |
|---|---|
| 25 | OTDEL0600 |
| 20 | OTDL0600 |
| 13 | OTDS0600 |




| L | CODE |
|---|---|
| 29 | OMDL0500 |
| 23 | OMDS0500 |
| 17 | OMDSS0500 |





| L | CODE |
|---|---|
| 32 | OMDEL0600 |
| 27 | OMDL0600 |
| 22 | OMDS0600 |

# Bite Impression KIT

**CODE**

BITE-KIT



- Bite Impression Coping: 20 Types.
- Scan Body: 10 Types (Regular & C Sold Separately).
- Coping Driver: 2 Types.
- Pinpoint: 1 Type.

## Bite Impression Coping



- Packing Unit: Bite Impression Coping Body + Guide Pin
- A component for Fixture Level impression taking.
- Allows for simultaneous impression and bite registration.
- Confirm implant seating by checking the screw and post height.
- Can also be used as a Scan Body; library available.
- Outer Driver: Short (OBICDSP), Long (OBICDLP).

**Size Guide**

- Guide Pin: Regular (Non-colored), Mini (Brown)
- Body: Color coded by G/H 2mm (Non), 3mm (Yellow), 4mm (Purple), 6mm (Blue), 8mm (Non).

### • Mini

| D | TYPE\G/H | 2.0 | 3.0 | 4.0 | 6.0 | 8.0 |
|---|---|---|---|---|---|---|
| Ø 4.0 | Hex | OBICM4002P | OBICM4003P | OBICM4004P | OBICM4006P | OBICM4008P |
| Ø 4.5 | Hex | OBICM4502P | OBICM4503P | OBICM4504P | OBICM4506P | OBICM4508P |

### • Regular

| D | TYPE\G/H | 2.0 | 3.0 | 4.0 | 6.0 | 8.0 |
|---|---|---|---|---|---|---|
| Ø 4.0 | Hex | OBICR4002P | OBICR4003P | OBICR4004P | OBICR4006P | OBICR4008P |
| Ø 4.5 | Hex | OBICR4502P | OBICR4503P | OBICR4504P | OBICR4506P | OBICR4508P |
| Ø 5.0 | Hex | OBICR5002P | OBICR5003P | OBICR5004P | OBICR5006P | OBICR5008P |
| Ø 6.0 | Hex | OBICR6002P | OBICR6003P | OBICR6004P | OBICR6006P | OBICR6008P |

## Bite Impression Coping Scanbody



- Packing Unit: Scan Body + Guide Pin.
- A Scan Body for fabricating Titanium Custom Abutments.
- For Model Scan: Long (15mm).
- For Intra-Oral Scan: Short (13mm).
- Hand-tighten using a 1.2 Hex Driver.

**Size Guide**

- Guide Pin: Regular (Non-colored), Mini (Brown)
- Body: Color coded by G/H 2mm (Non), 3mm (Yellow), 4mm (Purple), 6mm (Blue), 8mm (Non).

### • Mini

| D | TYPE\G/H | 2.0 | 3.0 | 4.0 | 6.0 | 8.0 |
|---|---|---|---|---|---|---|
| Ø 4.0 | Hex | OBICM4002SBP | OBICM4003SBP | OBICM4004SBP | OBICM4006SBP | OBICM4008SBP |

### • Regular


| D | TYPE\G/H | 2.0 | 3.0 | 4.0 | 6.0 | 8.0 |
|---|---|---|---|---|---|---|
| Ø 4.0 | Hex | OBICR4002SBP | OBICR4003SBP | OBICR4004SBP | OBICR4006SBP | OBICR4008SBP |
| Ø 4.5 | Hex | OBICR4502SBP | OBICR4503SBP | OBICR4504SBP | OBICR4506SBP | OBICR4508SBP |
| Ø 5.0 | Hex | OBICR5002SBP | OBICR5003SBP | OBICR5004SBP | OBICR5006SBP | OBICR5008SBP |

# Point Full S.S.T KIT

**CODE**

71105_DTSD

- Safe Sinus Trephine KIT.
- Sinus (Crestal + Lateral) Approach.





**CRESTAL REAMER**
Ø2.0 Ø3.3 Ø3.8 Ø4.2



**LATERAL DIAMOND REAMER Ø4.0/6.0**



**INITIAL DRILL Ø2.3/3.2**



**TISSUE PUNCH Ø4.0**



**AQUA LIFT TAP Ø2.8/3.8**



**SIDE CUTTER DRILL**



**AERO DRILL Ø3.2/3.8**



**REMOVER**



**SPRING STOPPER**



**DEPTH GAUGE**
DG-02_DTSD

**SINUS STOPPER**



3mm 4mm 5mm 6mm 7mm 8mm
(3mm) SCA-ST-03_DTSD (4mm) SCA-ST-04_DTSD
(5mm) SCA-ST-05_DTSD (6mm) SCA-ST-06_DTSD
(7mm) SCA-ST-07_DTSD (8mm) SCA-ST-08_DTSD

**Aero Drill Stopper**



3mm 4mm 5mm 6mm 7mm 8mm
(3mm) AD-ST-03_DTSD (4mm) AD-ST-04_DTSD
(5mm) AD-ST-05_DTSD (6mm) AD-ST-06_DTSD
(7mm) AD-ST-07_DTSD (8mm) AD-ST-08_DTSD



**BONE PUSHER**
BP-01_DTSD



**SINUS CURETS 5**
7095_DTSD



**SINUS CURETS 4**
7094_DTSD



**SINUS CURETS 3**
7093_DTSD



**SINUS CURETS 2**
7092_DTSD



**SINUS CURETS 1**
7091_DTSD



**AQUA PUMP SYRINGE**
APS-05_DTSD

# Point Simple S.S.T KIT

**CODE**

7130_DT3D

- Safe Sinus Trephine KIT
- Sinus (Crestal) Approach.




**SST BUR**
**Ø3.2**
For sinus floor elevation.
AD-32_DTSD

**INITIAL DRILL**
**Ø3.2**
For marking the ideal position during implant placement.
OSD-32T_DTSD

**AQUA TIP**
**Ø2.8/Ø3.8**
For hydraulically elevating the sinus membrane
ALT-28_DTSD

**SST BUR**
**Ø3.8**
For sinus floor elevation
AD-38_DTSD

**SST O-RING**
For replacing the inverted O-ring

**SST REMOVER**
For separating the drill part from the Shank
ADR_DTSD


**SST BUR STOPPER**
For elevation crest access during sinus approach.
AD-ST_DTSD

**TISSUE PUNCH**
**Ø3.8**
For gingival tissue removal, recommended for Flapless cases
6302_DTSD

# Point S.O.S KIT

**CODE**

7042_DTSD






SMALL MEDIUM

**IMPLANT REMOVER**

IRS-01S_DTSD
IRS-01M_DTSD



3.9/4.9 4.3/5.3 4.7/5.7 5.2/6.2 6.1/7.1

**EASY REMOVER_CLAW DRIVER**

FR-35_DTSD FR-50_DTSD
FR-40_DTSD FR-60_DTSD
FR-45_DTSD



**HAND/STRAUMAN ADAPTER**

HAS-01_DTSD



**CLAW DRIVER**

CD-01_DTSD



**SCREW DRILL**

SCD-01_DTSD



**TAP DRIVER**

TD-01_DTSD




**SCALER HANDPIECE**

SCR-01_DTSD
SCR-02_DTSD



ACTIVE RP | EXTERNAL 2.7 HEX | 8° MORSE TAPER(SMI) | SUBMERGED 2.5 HEX(SMII) | TSV 3.7/4.1 | CAMLOG TURBO-IN-TURBO

**SCREW GUIDE**

SG-02_DTSD
SG-03_DTSD
SG-04_DTSD
SG-05_DTSD
SG-06_DTSD
SG-12_DTSD
*Sold separately



**GUIDE WRENCH**

Used for fixing the screw guide to the implant (available separately).

# Bone Expander KIT

**CODE**

7030

- The expander's screw has a reverse-thread design, allowing it to accurately push and expand bone.
- Uses a pushing screw method instead of a bone-cutting method, enabling precise bone expansion surgery.





/ Reverse-thread design
/ BONE Trimmer
/ Color coding
/ Various sizes

Basic A TYPE & Master B TYPE



# Sinus Crestal Approach KIT

**CODE**

7020

- A kit for approaching and elevating the sinus membrane via the crestal area.
- Safe elevation is possible using Reamers and Diamond Burs.
- Hydraulically elevates the sinus membrane using an Aqua tip.
- Precisely and reliably designed Sinus Reamers.
- A simple hydraulic sinus membrane elevation system.
- Comprises tools for a fast and safe sinus elevation procedure.



# Sinus Lateral Approach KIT

**CODE**

7090

- A kit for approaching and elevating the sinus membrane via the lateral area.
- Allows safe and fast window opening with various Lateral Reamers.
- Comprehensive set of LATERAL REAMERS.
- Specialized design for easy accessibility.
- Reduces surgery time for LATERAL APPROACH.





# Implant & Screw Remover KIT(s)

**CODE**

7051

- Allows easy removal of a fractured abutment screw from inside the implant or the implant itself using a single kit.
- Includes Screw Guide and Screw Drill for easy and safe screw removal.
- One kit for removing fractured/damaged implants and abutment screws.
- Integrated system of Implant & Screw remover.
- Precise drilling using the Screw Guide protects the internal implant threads.



# Simple Screw Remover KIT

**CODE**

7052



- Used to directly remove a fractured abutment screw from inside the implant.
- A cost-effective and economical Mini Kit containing only essential items.
- Simple configuration for easier and more intuitive use.



Screw Driver CASE　Screw Drill　Top Driver CASE　Screw Drill

# Clinical Case Report 1
# Point UV Implant

Patient Information: Male, 57 years old

## Multi-staged Alveolar Ridge Splitting Technique


The patient presented with an inability to masticate and compromised esthetic function in the oral cavity. Implants were planned at multiple sites (#11, #13, #14, #16, #17, #18, #21, #23, #24, #25, #34, #35, #36, #44, #45, #46, #47).
To restore function, temporary implants were placed simultaneously with ridge splitting, allowing early functional recovery. Implant placement was performed after two stages of ridge splitting to compensate for insufficient alveolar bone width.
In the maxilla, implant placement was completed following two rounds of ridge splitting. Postoperatively, sufficient alveolar bone coverage over the cover screws was confirmed, allowing progression to the second surgical stage. Final treatment was completed with screw-retained prostheses placed in both the anterior and posterior regions.
At the 1-year follow-up, stable functional and esthetic outcomes were confirmed.



**First visit**
Loss of basic dental function, recurring multiple implant treatments.



**First visit**
As a result of CT section analysis, a lot of treatment is needed. With a knife-shaped narrow alveolar ridge in the maxilla implantation area.



**Primary ridge splitting before maxillary implant placement**
In the case of the maxilla, the bone density is low, so it is easily expanded only with a spreading chisel after cutting the crestal with piezo surgery. Temporary implant placed to prevent recurrence. Perform bone grafting of the expanded hollow complex. To minimize irritation at the implantation site, supporting implants were placed in the anterior and posterior regions.

# Clinical Case Report 1
# Point UV Implant

Patient information: Male, 57 years old





**After primary alveolar ridge splitting**
Temporary implant & support implant placement during denture function protects the DBM area and increases denture retention.





**2nd alveolar ridge splitting & implant placement**
6 months after the first alveolar ridge splitting. Implant placement while performing maxillary secondary ridge splitting.
Maintaining the supporting implant function using THA by increasing the initial fixation of the anterior region. Maintaining function after maxillary temporary denture correction.





**2nd surgery 6 months after maxillary implant placement**
Check that a sufficient amount of alveolar bone covered.
After fixing the abutment at the same time, the function is improved by installing a provisional restoration.





**End of treatment**
Anterior & posterior teeth screw type prosthetics completed. The treatment period was 1 year and 9 months.





**1 year F/U**
Functional and aesthetic stability.

# Clinical Case Report 2
# Point UV Implant + Point NAVI

Clinical Case 3
(#33,#34,#35,#37,#43,#44,#46,#47)

## 1. Before the surgery (X-ray, Clinical Picture)





## 2. Surgical Procedures



2-1) Extraction



2-2) Designed and customized surgical guiding template is manufactured and installed.



2-3) Anchored






The oral scan data and the CT data are combined to determine the optimal surgical angle and optimal placement position after mock surgery on the computer considering the position of the neural tube, occlusion, and bone tissue.

## 3. X-ray Picture after the surgery





The implant was placed at the appropriate position and depth as planned on the day of surgery.

## 4. Mounting the final prosthesis





The final prosthesis was fabricated and mounted using a crown. After mounting, the same normal occlusion as the normal teeth was confirmed.


## 5. Conclusion

In this case with Point Navigation system, it was able to secure an accurate implant position on the bone that was not enough, and it was possible to place the implant accurately at the positions # 37 and # 47 where other companies hard to access. Osseointegration was achieved without inflammatory diagnosis.

# Clinical Case Report 3
# Point UV Implant + Point NAVI

Clinical Case 4
(#13,#14,#16,#17,#23,#24,#26,#27)

## 1. Before the surgery (X-ray, Clinical Picture)




## 2. Surgical Procedures

2-1)




2-2)



Based on the CT scan and the oral scan image, the personal navigation device is made considering the establishment of the implantation route and the balance and the occlusion with the remaining teeth.


## 3. X-ray Picture after the surgery





The implant was placed at the appropriate position and depth as planned on the day of surgery.

## 3. X-ray Picture after the surgery




The final prosthesis was fabricated and mounted using a crown. After mounting, the same normal occlusion as the normal teeth was confirmed.

## 5. Conclusion

In this case with Point Navigation system, it was able to secure an accurate implant position on the bone that was not enough, and it was possible to place the implant accurately at the positions # 37 and # 47 where other companies hard to access. Osseointegration was achieved without inflammatory diagnosis.

# Clinical Case Report 4
# Point UV Implant

Patient Information: Female, 52 years old

## Multi-staged Alveolar Ridge Splitting Technique


The patient presented with severe esthetic and functional disorders affecting both the upper and lower jaws. Clinical examination revealed a Class III intermaxillary relationship with crossbite and insufficient alveolar bone volume for implant placement. Radiographic and clinical evaluation showed extremely narrow alveolar bone width in both sides of the mandible. On the right mandible, the alveolar bone was located very close to the inferior alveolar nerve, accompanied by severe vertical bone defects. The maxilla also exhibited insufficient horizontal bone width, although bone density was relatively favorable. Implants were planned and placed at sites #14, #16, #17, #24, #26, #27, #36, #37, #45, #46, #47.





**First visit**
Class III intermaxillary / crossbite / Mandibular alveolar bone is shallow in appearance.





#47 #46 #45 #44 #34 #35 #36 #37

**First visit**
Alveolar bone condition with insufficient bone width required for implant placement. Both sides of the mandible are very narrow, and the right mandible is very close to the inferior alveolar nerve, so the vertical bone defect is also serious.



**Maxilla**
Severe esthetic and functional problems of the patient. Immediate loading after maxillary implantation with ridge splitting.



**Mandible**
Primary ridge splitting (gap, expanded both lingually). Autogenous bone grafting between the enlarged alveolar bone to raise the vertical height.



**Mandible**
Severe bone fragment separation on the lower left side required light repositioning, followed by second ridge splitting and implant placement after 6–7 weeks.

# Clinical Case Report 4
# Point UV Implant

Patient Information: Female, 52 years old






**7 weeks after 1st Ridge splitting**
When the flap was opened, the fracture was stable, but the bone width was still narrow.
After the 2nd ridge splitting (both sides of the buccal and lingual), implantation and autogenous bone grafting.





**Right mandibular 2nd ridge splitting 9 weeks later**
The vertical dimension and bone width were still insufficient for implant placement, so after the 2nd ridge splitting, fixation with autogenous bone fragments, bone grafting, and non-absorbable membrane were used to cover the GBR area (to increase the vertical dimension)





**Right mandibular 2nd ridge splitting 4 months later**
Implant placement after 3rd ridge splitting to solve the condition of insufficient bone width.





**Mandibular Bilateral 2nd OP**
Sufficient amount of bone width regeneration confirmed on both sides.
In the case of the lower right side, a 1mm cover screw was used to further stabilize the marginal bone.





**1 year 4 months F/U**
stable alveolar bone condition.

# Clinical Case Report 5
# Point UV Implant

## Early loading at 1 week after mandibular posterior placement

A 83 - year - old female patient with history of implant overdenture in the maxillary and implant treatment in the mandibular at other dental clinic, she visited for reconstruction of uncomfortable implant dentures and mandibular implant retreatment. In the former dental clinic, two implants were placed on the right side of the mandible and three implants were placed on the left side, while the right side was compressing the mandibular nerve, and the two implants on the left side were removed with discomfort and pain sustained after implantation (Fig. 9). The maxillary cancellous defects were soft tissue laceration due to the superficial denture fissure and hypertrophic gingiva were observed, and the treatment was completed without adjustment of the denture base.
As a result of panoramic radiography and CT scan, the mandibular right side had insufficient space between the alveolar bone and the neural tube, so that first, the implant was placed on the left side to form the posterior occlusion. Vertical bone was appropriate, but the point width was somewhat narrow, so Point UV Active Ø4.0 * 10 mm was placed for # 35 and Point UV Active Ø4.0 * 8.5 mm fixture was placed for # 36 slightly deep without the exposed part to avoid bone graft, (Fig. 10). The bone quality confirmed when drilling was close to type 2 in # 35 and slightly weaker in # 36. The final setting torque was 45Ncm at the final placed depth of the # 35 fixture and 30Ncm at the final placed depth of the # 36 fixture. Healing abutments (Point Implant Co.) ware connected and sutured without any bone grafting, and a transfer impression was made after 3 days of suturing. After one week of operation, the suture was removed and the premature loading was performed with a temporary screw made of acrylic resin and a temporary implant abutment (Point Implant Co.). The temporary tooth was connected with the newly installed fixture and the previously mounted final fixture to achieve the fixation effect. The ISQ value increased in both fixtures after one week after the placement, and decreased slightly after the early loading. However, it increased again after one week or two weeks (Table 4). Two fixtures showed a stable healing pattern. After 2 months of placement (7 weeks of early loading), the final prosthesis was completed with a metal cyline porcelain tube connected to a 30Ncm of a standing abutment (Point Implant Co.)(Fig 12).



**Fig. 9. Case #3.**
Initial panoramic radiograph.



**Fig. 10. Case #3.**
After fixture placement.



**Fig. 11. Case #3.**
Three weeks after early loading
(4 weeks after fixture insertion).

## Table 4. ISQ value change during early loading (Case #3)

| | 1 st.Op. | 1 week | 2 weeks | 3 weeks | 4 weeks | 2 month |
|---|---|---|---|---|---|---|
| #35 | 79 | 85 | 81 | 79 | 80 | 82 |
| #36 | 70 | 71 | 68 | 73 | 74 | 75 |





**Fig. 12. Case #3.**
Final prosthesis.
(A) Abutment connection (2 months after fixture insertion).
(B) Three months after fixture installation.

# Clinical Case Report 6
# Point UV Implant

## Implant overdenture early loading after a week of implant placement

A 77 year old female patient who had undergone upper and lower dentures three years ago, but continued to complain of discomfort due to loose mandibular denture. As a result of examining the denture, the maxillary denture was good in retention, and the occlusal relationship of the maxillary and mandibular dentures was stable. In the case of the mandibular denture, there was a problem that the holding force of the mandibular denture was a problem, and two implants were placed in the mandibe and attached to the existing denture. (Fig. 13)
Implants were placed in both canine locations (#33 Point UV Active Ø4.0 × 10, #43 Point UV Active Ø4.0 × 11.5).
When drilling, the bone density was identified as type 2 and the insertion torque was recorded at 40Ncm in both final placed depths (Fig. 14).
After the healing abutments were connected and sutured, only the area where the healing zone was located was removed and used by the patient. 1 week later, the suture was removed and the Locator attachment (LOCATOR, Zest Anchors LLC, USA) was fixed to the implant at 20Ncm and the denture cap (Zest Anchors LLC) the black processing male (Zest Anchors LLC) (Fig. 15).
As shown in Table 5, the increase in ISQ values after one week of implantation was very impressive, and the increased values remained very stable even after early loading. After 4 weeks of placement, the locator attachment was tightened to 30Ncm and the black processing male was replaced with a blue LOCATOR male (Zest Anchors LLC).



**Fig. 13. Case #4.**
Initial panoramic radiographic



**Fig. 14. Case #4.**
Two fixtures were installed in canine area of mandible.







**Fig. 15. Case #4.**
(A) Two LOCATOR attachments were connected to fixtures. Denture caps with black processing male were set on the Locator attachments.
(B) Relief of denture base.
(C) Denture caps were embedded in old denture with auto-polymerizing resin (Pattern Resin LS, GC, USA) in direct technique.

## Table 5. ISQ value changes during early loading (Case #4)

| | 1 st.Op. | 1 week (loading) | 2 weeks | 3 weeks | 4 weeks |
|---|---|---|---|---|---|
| #33 | 72 | 80 | 83 | 85 | 86 |
| #43 | 81 | 86 | 86 | 85 | 85 |

# Clinical Case Report 7
# Point UV Implant

## Multi-staged alveolar ridge splitting technique

A 57 year old male (born in 1965) presented with impaired masticatory function and compromised esthetics in the oral cavity, involving teeth #11, #13, #14, #16, #17, #18, #21, #23, #24, #25, #34, #35, #36, #44, #45, #46, and #47.
Ridge splitting was performed twice in the maxilla, and temporary implants were placed simultaneously to facilitate functional recovery.
Following implant placement, sufficient alveolar bone coverage over the cover screws was confirmed, allowing progression to the second-stage surgery.
Definitive screw-retained prostheses were installed in both the anterior and posterior regions. At the 1-year follow-up, stable functional performance and satisfactory esthetic outcomes were confirmed.





At the first visit, the patient presented with a loss of basic dental function, requiring multiple implant treatments.
CT section analysis revealed a knife-edge shaped narrow alveolar ridge in the maxillary implant site, noticing the need for extensive treatment.





**Primary Ridge Splitting**
Primary alveolar ridge splitting was performed prior to maxillary implant placement.
Due to the low bone density of the maxilla, ridge expansion was achieved using a screeding chisel after crestal incision with piezo surgery. Temporary implants were placed to prevent recase of the expanded ridge.
Porcine bone grafting was performed in the expanded hollow area. To minimize irritation at the implant site, supporting implants were placed in both the anterior and posterior regions.
**Post-Primary Ridge Splitting**
Temporary and supporting implants were maintained during denture function, protecting the GBR area and improving denture retention.
Compared to the initial examination, a significant expansion of the alveolar bone was observed.
**Secondary Ridge Splitting & Implant Placement**
Six months after the primary ridge splitting, secondary maxillary ridge splitting and implant placement were performed simultaneously.
Initial fixation in the anterior region was enhanced, and supporting implant function was maintained using THA.
Functional stability was preserved following correction of the maxillary temporary denture.




**End of Treatment**
Definitive screw-retained prostheses were completed in both the anterior and posterior regions.
The total treatment period was 1 year and 6 months.

# Clinical Case Report 8
# Point Short Implant

M 60Y 2024.02.06



A comprehensive rehabilitation following severe bone resorption and previous implant failure, utilizing Point UV implants in combination with Short UV implants to address areas of low residual bone height and close proximity to the inferior alveolar nerve.

**Fig. 1** Panoramic radiograph and intraoral photographs

**Surgical Procedure**



**Fig. 2** Replacement of failed implants in the mandible.



**Fig. 3** Placement of multi-unit abutments and immediate loading with a provisional prosthesis

2024.02.13 2024.04.03



**Fig. 4** Implant placement in the maxilla and immediate loading with a provisional prosthesis.

2025.09.09




**Fig. 5** Placement of Point Short UV Implants in close proximity to the inferior alveolar nerve at sites #37, #46, and #47.

**Post-Operative Documentation: Final panoramic radiograph and intraoral photographs.**

2025.12.18




**Conclusion**



This case demonstrates the successful application of a combined implant protocol for a complex full-mouth rehabilitation. By strategically utilizing Point UV implants for standard sites and Short UV implants in areas of severe anatomical limitation—specifically where low residual bone height and extreme proximity to the inferior alveolar nerve precluded conventional implant placement—a predictable and biomechanically sound foundation was established.

# Case Report 9
# Point UV Implant

Patient Information: Male, 65 years old

## Point UV Implant & Point Guide

**Preoperative Condition (X-ray Images)**

Preoperative radiographic images were taken to evaluate the patient's oral condition and treatment plan prior to surgery.




**Surgical Procedure**

Based on the confirmed surgical plan, a surgical guide was prepared, and the implant fixture sizes selected during treatment planning were verified and pre-treated using a UV activator in advance.





Precise drilling was performed with sufficient saline irrigation. During UV implant placement, a **Smart Fixture Driver** was used to adjust the insertion depth and hex orientation to enable immediate loading.






**Surgical Outcome (Immediate Function)**

On the day of surgery, a pre-fabricated **PMMA bridge** and **customized abutments** were connected to provisionally restore the collapsed vertical dimension.

During follow-up visits, the vertical dimension was gradually adjusted through continuous communication with the patient.





**Final Prosthesis Placement**

The patient's vertical dimension was evaluated and applied using a **temporary PMMA bridge**. Based on this assessment, the occlusion of the final prosthesis was carefully verified to ensure a stable and normal bite.





# Case Report 10
# Point UV Implant

Patient Information: Female, 60 years old

## | Point UV Implant

**Preoperative Condition (X-ray Images)**




**C.C.) Bilateral maxillary posterior implant failure at another dental clinic.**
**#14–17 :** Failed implants placed at another clinic; extraction performed several months prior.
**#24–26 :** Failed implants placed at another clinic; extraction performed two weeks prior to surgery.
**Tx. Plan)**
**#13 :** Immediate implant placement following tooth extraction.
**#14, 15, 16, 17, 24, 25, 26 :** Placement of Point UV Implants with bilateral maxillary sinus elevation.

**Surgical Procedure**
Fixture Installation: Point UV Implant
Sinus Elevation: Bilateral lateral window approach using CGF and PRF (PointGen®)





**Postoperative Panoramic Evaluation (Early Loading)**
Early loading was initiated 10 days after surgery on the right maxilla.





At site #13, the implant was placed immediately following tooth extraction.
Although implants at sites #14, 15, 16, 17, 24, 25, and 26 were placed in bone defect areas resulting from failed implants from another manufacturer, stable initial responses were observed.
As a result, temporary PMMA bridges were able to be installed 10 days after implant placement at sites #13, 14, 15, and 17.

**ISQ values on the day of surgery vs. 1 week postoperatively**

| | #13 | #14 | #15 | #17 |
|---|---|---|---|---|
| Day of Surgery | 82 | 81 | 78 | 68 |
| 1Week After | 87 | 84 | 80 | 71 |

**ISQ values on the day of surgery vs. 3 weeks postoperatively**

| | 24 | 25 | 26 |
|---|---|---|---|
| Day of Surgery | 80 | 93 | 74 |
| 3Weeks After | 81 | 90 | 84 |

**Final Prosthesis Placement**
This case demonstrates successful clinical outcomes using Point UV implants in sites with previous implant failures.
Despite the unfavorable conditions of failed implant sites, early loading was feasible, and stable osseointegration was achieved.
Subsequent prosthetic treatment progressed successfully without differences compared to conventional sites, and the results have been well maintained both functionally and esthetically to date.







# Case Report 11
# Point UV Implant

Patient Information: Male, 51 years old

## Alveolar Ridge Splitting

The patient presented with severe generalized periodontitis affecting both the maxillary and mandibular dentition, requiring extraction of all remaining teeth. The involved teeth were #12, #13, #14, #15, #16, #17, #21, #23, #25, #26, #27, #33, #34, #35, #37, #43, #44, #45, and #47. Immediate loading was achievable in the mandible following full-mouth extraction. In contrast, the maxilla required extensive treatment, including maxillary sinus bone grafting and guided bone regeneration (GBR), to secure sufficient bone volume for implant placement.





**First Visit**
Diagnosis of severe periodontitis involving all maxillary and mandibular teeth.
**First Mandibular Implant Placement**
Extraction of all mandibular teeth and implant placement for partial functional recovery.
Temporary implants were placed at maxillary #12, with teeth #13, #21, and #25 retained temporarily, followed by delivery of a maxillary provisional prosthesis.





**Pre-Implant Evaluation**
Evaluation of reconstructed alveolar bone and confirmation of successful 6-month GBR healing in the maxillary anterior region.
**Maxillary Implant Placement & Immediate Loading**
Implant placement in the maxillary anterior region with simultaneous ridge splitting.
Immediate loading was performed following extraction of the remaining maxillary anterior teeth.
**Outcome**
The alveolar bone demonstrated stable integration, with satisfactory functional and esthetic outcomes maintained after treatment.





2020.04.23 / 1st prosthetic treatment
#34, 35, 36, 37 #44, 45, 46, 47

2020.05.14 / 2nd prosthetic treatment
#33-43 #16, 17, 26, 27

2020.07.14 / 3rd prosthetic treatment
#15-25





**2021. 02. 17 / 1 year 10 months F/U**
Aesthetically and functionally stable in the state of regular checkup for about 2 years

# Case Report 12
# Point UV Implant

Patient Information: Female, 64 years old

## | Splitting & Long term stability

This case involved a patient with chronic periodontitis affecting multiple teeth, resulting in pain and masticatory dysfunction. Implants were placed at sites #14, #16, #17, #24, #26, #27, #36, #37, #45, #46, and #47 using ridge splitting and guided bone regeneration (GBR) techniques. Despite alveolar bone resorption caused by periodontitis and involvement of the maxillary sinus, stable bone ossification, osseointegration, and reliable occlusal stability were successfully achieved around the implants.



First Visit

The patient visited the clinic with pain and masticatory dysfunction caused by chronic periodontitis affecting multiple teeth. Clinical and radiographic evaluation revealed reduced bone width and low bone density in the maxillary left alveolar bone.



Implant Placement (2018.08.13 / 2018.10.04)

Implants were placed at sites #14, #16, #17, #18, #45, #46, #47, followed by additional implant placement at #24, #26, #27, #36, and #37 using ridge splitting and GBR techniques to compensate for insufficient bone volume. Immediate loading with a temporary prosthesis was applied at #45, #46, and #47, and temporary implants were connected with natural teeth to restore function at an early stage.

Prosthetic Treatment (2019.02.12 – 2019.04.26)

Final prosthetic treatment was completed in stages.




2019. 02. 12 / 1st prosthetic treatment
#14, #15, #16, #17, #45, #46, #47
Completion of final prosthetic treatment.

2019. 04. 18 / 2nd prosthetic treatment
#24, #26, #27, #36, #37
Implant final prosthetic treatment completed.

2019. 04. 26 / final prosthetic treatment
Maxillary anterior prosthetic treatment completed.

Follow-up Evaluation




2021. 02. 19 / 1 year 10 months F/U
It shows a stable appearance in the marginal bone around the stable implant and shows gradually increasing bone density in the implanted alveolar bone area.

2022. 07. 29 / 3 years 3 months F/U
CT scans of #26 and #27 where ridge splitting was performed showed bone width expansion and bone density increase.

# What is Point NAVI?

This is a cutting-edge program that allows you to place the implant in the most ideal position while checking the shape of the final prosthesis by building a precise 3D image.




## Digital point navigator to preview implant treatment



**1. Highest accuracy and safety**

Safe operation by predicting risk factors more precisely using 3D digital guide.



**2. No incision procedure**


Reduce pain and speed up recovery through incision-free treatment by digital navigation guide.



**3. One Day Implant**

One day implant that can be used from implant placement to prosthesis just like your teeth on the day of surgery.

## Point NAVI Workflow




## Point NAVI Advantages

Point Navi guides the optimal implant surgery in a simple and safe manner through fast surgical guide production and accurate process without errors as all processes are digitalized from start to finish.



**1. Highest accuracy and safety**

Predicting risk factors more precisely and by using 3D digital guides obtain safety of operation.



**2. Flapless Surgery**

Reduce pain and speed up recovery through incision-free surgery by digital navigation guide.



**3. One Day Implant**

One-Day implant that can be used from implant placement to prosthesis just like your teeth on the same day of surgery.

## Point Guide System



Point Guide S/W | Surgical Guide | Point Guide KIT | Immediate Loading



## • Point 3D Combi 900S

| | |
|---|---|
| Model | Point 3D Combi 900S |
| Modality | Panoramic, CBCT, Cephalometric |
| F.O.V | (Single) 12X9, 10X9<br>(Multi) 16X15, 16X10, 10X10 |
| Unit dimension | 1,873mm x 1,150mm x 2,280mm |
| Frequency ranges | 0 ~ 20kHz |
| X-ray Tube | D-054SB (Focal Spot 0.5mm) |
| kVp / mA | 60 ~ 90 kVp / 4 ~ 10mA |
| Input voltage | 100 ~ 120V / 220 ~ 240V, 50/60Hz |
| Weight | 184 kg |

## • Point 800S HD 3D Plus

| | |
|---|---|
| Model | Point 800S HD 3D Plus |
| Modality | Panoramic, CBCT, Cephalometric |
| F.O.V | (Single) 12X9 or 10X9 |
| Panorama sensor | Independent and Separate |
| Unit dimension | 1,854mm x 1,150mm x 2,290mm |
| Frequency ranges | 0 ~ 20kHz |
| X-ray Tube | D-054SB (Focal Spot 0.5mm) |
| kVp / mA | 50 ~ 100 kvp / 4~16mA |
| Input voltage | 100 ~ 120V / 220 ~ 240V, 50/60Hz |
| Weight | 190kg |



## • Point 3D Combi 500S/C

| | |
|---|---|
| Modal | Point 3D Combi 500S/C |
| Modality | Panoramic, CBCT, Cephalometric |
| F.O.V | (Single) 12X9, 10X9<br>(Multi) 16X15, 16X10, 10X10 |
| Unit dimension | 1,858mm x 1,190mm x 2,290 mm |
| Frequency ranges | 0 ~ 20kHz |
| X-ray Tube | D-054SB (Focal Spot 0.5mm) |
| kVp / mA | 60 ~ 90kVp / 4 ~ 10mA |
| Input voltage | 100 ~ 120V / 220 ~ 240V 50/60Hz |
| Weight | 190 kg |

## • Point 200 HD

| | |
|---|---|
| Modal | Point 200 HD |
| Modality | Panoramic Only |
| Unit dimension | 680mm x 790mm x 2,310 mm |
| Frequency ranges | 0 ~ 20kHz |
| X-ray Tube | D-054SB (Focal Spot 0.5mm) |
| kVp / mA | 60 ~ 90kVp / 4 ~ 10mA |
| Input voltage | 100 ~ 120V / 220 ~ 240V 50/60Hz |
| Weight | 150 kg |



## • SK - Iα

| | |
|---|---|
| Elctrical rating | AC 220V, 60Hz, 1.2kVA |
| Hand Piece System | Air control system |
| Lifting Mechanism | Electro hydraulic motor |
| IChair Hcight | Max 690±20mm, Min 430±20mm |
| Chair Length | 1,900mm |
| Light | ≥ 16,000 lux |

## • SK - IIα

| | |
|---|---|
| Elctrical rating | AC 220V, 60Hz, 1.2kVA |
| Hand Piece System | Electric control system |
| Lifting Mechanism | Electro hydraulic motor |
| IChair Hcight | Max 690±20mm, Min 430±20mm |
| Chair Length | 1,900mm |
| Light | ≥ 16,000 lux |



- **SK - Iβ, IIβ (Equipped with the same control system)**

| | |
|---|---|
| Electrical rating | AC 220V, 60Hz, 1.2kVA |
| Hand Piece System | Refined base design |
| Lifting Mechanism | Electro-hydraulic motor |
| Chair Height | Max 690±20mm, Min 430±20mm |
| Chair Length | 1,900mm |
| Light | ≥ 16,000 lux |

- **Arte**

| | |
|---|---|
| Electrical rating | AC 220V, 60Hz, 1.2kVA |
| Hand Piece System | Electric control system |
| Lifting Mechanism | Electro-hydraulic motor |
| Chair Height | Max 690±20mm, Min 430±20mm |
| Chair Length | 1,900mm |
| Light | ≥ 16,000 lux |

# POINT Quantum Laser

## **Masterpiece of POINT,**
Experience the state-of-the-art **Quantum Laser.**

| | |
|---|---|
| Wavelength | Er:YAG 2940nm<br>Nd:YAG 1064nm / 1320nm |
| Maximum Output Power | Er:YAG 6W<br>Nd:YAG 10W / 8W |
| Maximum Pulse Energy | Er:YAG: 400 mJ<br>Nd:YAG: 600 mJ / 500 mJ / 1560 mJ/cm² |

| | |
|---|---|
| Maximum Pulse Repetition Rate | Er:YAG: 50 Hz<br>Nd:YAG: 50 Hz / 30 Hz |
| Pulse width | Er:YAG: 200 µs / 600 µs<br>Nd:YAG: 200 ~ 600 µs |
| Dimensions | 925(H) x 270(W) x 530(D) |
| Weight | 50kg |







## • Aqua ToTo

| | |
|---|---|
| Wavelength | Er:YAG 2940nm |
| Max. output power | Er:YAG 8W |
| Pulse Energy | Er:YAG 600mJ |
| Pulse Repetition Rate | 10 ~ 50Hz |
| Pulse width | Er:YAG : 200 μs |
| Dimensions | 270(W) x 430(D) x 780(H) |
| Weight | 41kg |

## • AnyBeam Cera

| | |
|---|---|
| Wavelength | Nd:YAG 1064nm / 1320nm |
| Max. output power | Nd:YAG 10W / 8W |
| Pulse Energy | 600mJ(1064nm), 500mJ(1320nm) |
| Pulse repetition rate | Nd:YAG 1, 5 ~ 50Hz<br>(Adjustable in 5 Hz increments) |
| Transmission Media | Optical fiber |
| Aiming beam | Red diode laser, 650 nm |
| Dimensions | 268(W) x 553(D) x 862(H) |
| Weight | 50kg |

Point Implant

Good implants. Be happy world!


**Head Office** +82-2-6480-0110

7F, 8F, 25F, Building C, GIDC Gwangmyeong Station, 43 Iljik-ro, Gwangmyeong-si, Gyeonggi-do, Republic of Korea

**Plant**

7 Hyeyeol-ro, Buk-gu, Busan, South Korea

**Tel**

**1600-2369**

www.pointimplant.com



Ver. 15.1